

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC092022AL\
 Data File : FC061578.D
 Signal(s) : FID1A.ch
 Acq On : 20 Sep 2022 15:57
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD3
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
 Quant Time: Sep 20 16:43:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 092022.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 20 16:42:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.573	3230728	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
12) S 1-chlorooctadecane (S...	13.012	2581862	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
Target Compounds			
1) T n-Nonane (C9)	3.320	2812275	20.000 ug/ml
2) T n-Decane (C10)	4.391	2854598	20.000 ug/ml
3) T A~Naphthalene (C11.7)	5.978	3133599	20.000 ug/ml
4) T n-Dodecane (C12)	6.412	2872389	20.000 ug/ml
5) T A~2-methylnaphthalene...	7.035	3101722	20.000 ug/ml
6) T n-Tetradecane (C14)	8.211	2881347	20.000 ug/ml
7) T n-Hexadecane (C16)	9.814	2926310	20.000 ug/ml
8) T n-Octadecane (C18)	11.257	2976249	20.000 ug/ml
10) T n-Eicosane (C20)	12.568	3006402	20.000 ug/ml
11) T n-Heneicosane (C21)	13.180	3030818	20.000 ug/ml
13) T n-Docosane (C22)	13.766	3031641	20.000 ug/ml
14) T n-Tetracosane (C24)	14.869	3025963	20.000 ug/ml
15) T n-Hexacosane (C26)	15.889	2997325	20.000 ug/ml
16) T n-Octacosane (C28)	16.839	2933387	20.000 ug/ml
17) T n-Tricontane (C30)	17.726	2895837	20.000 ug/ml
18) T n-Dotriacontane (C32)	18.558	2825513	20.000 ug/ml
19) T n-Tetratriacontane (C34)	19.342	2652586	20.000 ug/ml
20) T n-Hexatriacontane (C36)	20.081	2612876	20.000 ug/ml
21) T n-Octatriacontane (C38)	20.815	2450500	20.000 ug/ml
22) T n-Tetracontane (C40)	21.727	2296140	20.000 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC092022AL\
Data File : FC061578.D
Signal(s) : FID1A.ch
Acq On : 20 Sep 2022 15:57
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD3
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
Quant Time: Sep 20 16:43:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 092022.M
Quant Title : GC Extractables
QLast Update : Tue Sep 20 16:42:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 uL
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

