

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 09/22/2022
Supervised By :Ankita Jodhani 09/22/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC092022
Data File : FC061589.D
Signal(s) : FID1A.ch
Acq On : 21 Sep 2022 10:39
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 092022.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.328	3.300	3.362	BB	334630	2806079	87.01%	4.443%
2	4.395	4.365	4.445	BB	313170	2831466	87.79%	4.483%
3	5.978	5.947	6.014	BV	335847	3109592	96.42%	4.923%
4	6.412	6.374	6.447	BB	290592	2850402	88.38%	4.513%
5	7.035	6.990	7.062	BV	317373	3081792	95.56%	4.879%
6	8.210	8.172	8.242	BB	265010	2858283	88.63%	4.526%
7	9.813	9.739	9.857	BB	259478	2905338	90.08%	4.600%
8	11.256	11.195	11.309	BB	254137	2960761	91.80%	4.688%
9	11.571	11.505	11.620	BB	289933	3225110	100.00%	5.106%
10	12.565	12.497	12.605	BB	243981	2996151	92.90%	4.744%
11	13.010	12.937	13.049	BB	211157	2570270	79.70%	4.069%
12	13.177	13.109	13.222	BB	245771	3013482	93.44%	4.771%
13	13.763	13.699	13.819	BB	238299	3023334	93.74%	4.787%
14	14.867	14.790	14.914	BB	227375	3014051	93.46%	4.772%
15	15.888	15.832	15.929	BB	209974	2980395	92.41%	4.719%
16	16.837	16.759	16.895	BB	205451	2916679	90.44%	4.618%
17	17.724	17.575	17.769	BB	206791	2951581	91.52%	4.673%
18	18.556	18.489	18.622	BB	205728	3019698	93.63%	4.781%
19	19.340	19.250	19.405	BV	190728	2822347	87.51%	4.469%
20	20.078	20.030	20.110	BV	176258	2619441	81.22%	4.147%
21	20.811	20.752	20.898	M	136204	2509130	77.80%	3.973%
22	21.723	21.652	21.775	M	91461	2094084	64.93%	3.316%
Sum of corrected areas:						63159467		

Aliphatic EPH 092022.M Wed Sep 21 11:14:48 2022