

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091923\
 Data File : PP060016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2023 11:31
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 11:53:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 09:22:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.510	3.668	97659014	74533693	46.945	45.758
2)	SA Decachlor...	10.430	8.743	75319867	75604275	49.379	49.931
Target Compounds							
3)	L1 AR-1016-1	5.697	4.768	31761802	27990185	470.719	465.606
4)	L1 AR-1016-2	5.720	4.788	45246507	38596541	472.987	465.009
5)	L1 AR-1016-3	5.783	4.965	28408395	21340709	477.268	475.729
6)	L1 AR-1016-4	5.883	5.007	23301940	16273039	475.878	461.268
7)	L1 AR-1016-5	6.182	5.223	23950897	22495313	487.125	483.806
8)	L2 AR-1221-1	4.717	3.884	3951690	3243908	148.654	147.686
9)	L2 AR-1221-2	4.808	3.971	5438155	4078489	276.286	264.170
10)	L2 AR-1221-3	4.886	4.049	18808951	14824262	329.295	316.061
11)	L3 AR-1232-1	4.886	4.049	18808951	14824262	403.152	389.973
12)	L3 AR-1232-2	5.425	4.788	25545072	38596541	1005.762	1046.739
13)	L3 AR-1232-3	5.720	4.965	45246507	21340709	1047.292	1065.783
14)	L3 AR-1232-4	5.883	5.050	23301940	21369156	1054.249	1169.852
15)	L3 AR-1232-5	5.975	5.223	20971500	22495313	1143.150	1122.001
16)	L4 AR-1242-1	5.697	4.768	31761802	27990185	531.795	531.516
17)	L4 AR-1242-2	5.720	4.788	45246507	38596541	538.411	535.643
18)	L4 AR-1242-3	5.783	4.965	28408395	21340709	543.166	545.329
19)	L4 AR-1242-4	5.883	5.050	23301940	21369156	540.748	545.278
20)	L4 AR-1242-5	6.630	5.579	7155447	22660193	156.508	468.379 #
21)	L5 AR-1248-1	5.697	4.768	31761802	27990185	710.608	716.362
22)	L5 AR-1248-2	5.975	5.007	20971500	16273039	323.815	300.332
23)	L5 AR-1248-3	6.182	5.050	23950897	21369156	338.182	375.559
24)	L5 AR-1248-4	6.590	5.223	8120629	22495313	104.993	334.934 #
25)	L5 AR-1248-5	6.630	5.620	7155447	6170905	95.101	93.550
26)	L6 AR-1254-1	6.564	5.579	21317266	22660193	267.012	225.796
27)	L6 AR-1254-2	6.786	5.728	18233954	17684677	152.239	203.561 #
28)	L6 AR-1254-3	7.156	6.152	9011770	30509221	74.183	217.301 #
29)	L6 AR-1254-4	7.445	6.366	5934839	2294131	66.736	27.254 #
30)	L6 AR-1254-5	7.867	6.787	56362780	53760290	579.416	453.223
31)	L7 AR-1260-1	7.323	6.267	44062388	44116612	488.829	473.780
32)	L7 AR-1260-2	7.582	6.456	49514400	51075141	482.654	465.856
33)	L7 AR-1260-3	7.945	6.612	38481553	46377440	488.602	450.904
34)	L7 AR-1260-4	8.175	7.088	43726248	39317813	487.905	472.094
35)	L7 AR-1260-5	8.509	7.330	79499916	87024460	465.837	452.303
36)	L8 AR-1262-1	7.945	6.880	38481553	20955776	332.502	415.180
37)	L8 AR-1262-2	8.509	7.088	79499916	39317813	397.121	348.023
38)	L8 AR-1262-3	8.849	7.618	53608888	19294845	372.282	220.125 #
39)	L8 AR-1262-4	8.912f	7.681	36049493	60590596	317.459	371.024
40)	L8 AR-1262-5	9.624	8.181	23023779	19479965	308.618	260.974
41)	L9 AR-1268-1	8.849	7.618	53608888	19294845	213.787	74.110 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091923\
 Data File : PP060016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2023 11:31
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 11:53:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 19 09:22:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.912f	7.681	36049493	60590596	157.077	262.082 #
43) L9	AR-1268-3	9.175	7.893	1160014	1100834	5.879	5.392
44) L9	AR-1268-4	9.624	8.181	23023779	19479965	277.975	241.255
45) L9	AR-1268-5	10.065	8.479	6443604	6014460	10.646	9.522



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP091923\
Data File : PP060016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2023 11:31
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 11:53:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP091823.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 19 09:22:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.510 min
Delta R.T.: 0.000 min
Response: 97659014
Conc: 46.95 ng/ml

Instrument :
FID_C
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
Delta R.T.: -0.003 min
Response: 74533693
Conc: 45.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.430 min
Delta R.T.: -0.002 min
Response: 75319867
Conc: 49.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.743 min
Delta R.T.: -0.015 min
Response: 75604275
Conc: 49.93 ng/ml



#3 AR-1016-1


R.T.: 5.697 min
Delta R.T.: -0.002 min
Response: 31761802
Conc: 470.72 ng/ml

Instrument :
FID_C
ClientSampleId :




#3 AR-1016-1

R.T.: 4.768 min
Delta R.T.: -0.006 min
Response: 27990185
Conc: 465.61 ng/ml




#4 AR-1016-2

R.T.: 5.720 min
Delta R.T.: -0.002 min
Response: 45246507
Conc: 472.99 ng/ml



#4 AR-1016-2


R.T.: 4.788 min
Delta R.T.: -0.006 min
Response: 38596541
Conc: 465.01 ng/ml



#5 AR-1016-3


R.T.: 5.783 min
Delta R.T.: -0.002 min
Response: 28408395
Conc: 477.27 ng/ml

Instrument :
FID_C
ClientSampleId :




#5 AR-1016-3

R.T.: 4.965 min
Delta R.T.: -0.007 min
Response: 21340709
Conc: 475.73 ng/ml




#6 AR-1016-4

R.T.: 5.883 min
Delta R.T.: -0.002 min
Response: 23301940
Conc: 475.88 ng/ml



#6 AR-1016-4


R.T.: 5.007 min
Delta R.T.: -0.007 min
Response: 16273039
Conc: 461.27 ng/ml



#7 AR-1016-5


R.T.: 6.182 min
Delta R.T.: -0.003 min
Response: 23950897
Conc: 487.12 ng/ml

Instrument :
FID_C
ClientSampleId :




#7 AR-1016-5

R.T.: 5.223 min
Delta R.T.: -0.008 min
Response: 22495313
Conc: 483.81 ng/ml




#8 AR-1221-1

R.T.: 4.717 min
Delta R.T.: -0.004 min
Response: 3951690
Conc: 148.65 ng/ml



#8 AR-1221-1


R.T.: 3.884 min
Delta R.T.: -0.005 min
Response: 3243908
Conc: 147.69 ng/ml



#9 AR-1221-2


R.T.: 4.808 min
Delta R.T.: 0.000 min
Response: 5438155
Conc: 276.29 ng/ml

Instrument :
FID_C
ClientSampleId :




#9 AR-1221-2

R.T.: 3.971 min
Delta R.T.: -0.005 min
Response: 4078489
Conc: 264.17 ng/ml




#10 AR-1221-3

R.T.: 4.886 min
Delta R.T.: -0.001 min
Response: 18808951
Conc: 329.30 ng/ml



#10 AR-1221-3


R.T.: 4.049 min
Delta R.T.: -0.005 min
Response: 14824262
Conc: 316.06 ng/ml



#11 AR-1232-1


R.T.: 4.886 min
Delta R.T.: 0.000 min
Response: 18808951
Conc: 403.15 ng/ml

Instrument :
FID_C
ClientSampleId :




#11 AR-1232-1

R.T.: 4.049 min
Delta R.T.: -0.004 min
Response: 14824262
Conc: 389.97 ng/ml




#12 AR-1232-2

R.T.: 5.425 min
Delta R.T.: 0.000 min
Response: 25545072
Conc: 1005.76 ng/ml



#12 AR-1232-2


R.T.: 4.788 min
Delta R.T.: -0.005 min
Response: 38596541
Conc: 1046.74 ng/ml



#13 AR-1232-3


R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 45246507
Conc: 1047.29 ng/ml

Instrument :
FID_C
ClientSampleId :




#13 AR-1232-3

R.T.: 4.965 min
Delta R.T.: -0.005 min
Response: 21340709
Conc: 1065.78 ng/ml




#14 AR-1232-4

R.T.: 5.883 min
Delta R.T.: -0.002 min
Response: 23301940
Conc: 1054.25 ng/ml



#14 AR-1232-4


R.T.: 5.050 min
Delta R.T.: -0.007 min
Response: 21369156
Conc: 1169.85 ng/ml



#15 AR-1232-5


R.T.: 5.975 min
Delta R.T.: 0.000 min
Response: 20971500
Conc: 1143.15 ng/ml

Instrument :
FID_C
ClientSampleId :




#15 AR-1232-5

R.T.: 5.223 min
Delta R.T.: -0.007 min
Response: 22495313
Conc: 1122.00 ng/ml




#16 AR-1242-1

R.T.: 5.697 min
Delta R.T.: -0.002 min
Response: 31761802
Conc: 531.79 ng/ml



#16 AR-1242-1


R.T.: 4.768 min
Delta R.T.: -0.006 min
Response: 27990185
Conc: 531.52 ng/ml



#17 AR-1242-2


R.T.: 5.720 min
Delta R.T.: -0.002 min
Response: 45246507
Conc: 538.41 ng/ml

Instrument :
FID_C
ClientSampleId :




#17 AR-1242-2

R.T.: 4.788 min
Delta R.T.: -0.006 min
Response: 38596541
Conc: 535.64 ng/ml




#18 AR-1242-3

R.T.: 5.783 min
Delta R.T.: -0.002 min
Response: 28408395
Conc: 543.17 ng/ml



#18 AR-1242-3


R.T.: 4.965 min
Delta R.T.: -0.006 min
Response: 21340709
Conc: 545.33 ng/ml



#19 AR-1242-4


R.T.: 5.883 min
Delta R.T.: -0.002 min
Response: 23301940
Conc: 540.75 ng/ml

Instrument :
FID_C
ClientSampleId :




#19 AR-1242-4

R.T.: 5.050 min
Delta R.T.: -0.008 min
Response: 21369156
Conc: 545.28 ng/ml




#20 AR-1242-5

R.T.: 6.630 min
Delta R.T.: -0.003 min
Response: 7155447
Conc: 156.51 ng/ml



#20 AR-1242-5


R.T.: 5.579 min
Delta R.T.: -0.006 min
Response: 22660193
Conc: 468.38 ng/ml



#21 AR-1248-1


R.T.: 5.697 min
Delta R.T.: 0.000 min
Response: 31761802
Conc: 710.61 ng/ml

Instrument :
FID_C
ClientSampleId :




#21 AR-1248-1

R.T.: 4.768 min
Delta R.T.: -0.004 min
Response: 27990185
Conc: 716.36 ng/ml




#22 AR-1248-2

R.T.: 5.975 min
Delta R.T.: 0.000 min
Response: 20971500
Conc: 323.81 ng/ml



#22 AR-1248-2


R.T.: 5.007 min
Delta R.T.: -0.006 min
Response: 16273039
Conc: 300.33 ng/ml



#23 AR-1248-3


R.T.: 6.182 min
Delta R.T.: 0.000 min
Response: 23950897
Conc: 338.18 ng/ml

Instrument :
FID_C
ClientSampleId :




#23 AR-1248-3

R.T.: 5.050 min
Delta R.T.: -0.005 min
Response: 21369156
Conc: 375.56 ng/ml




#24 AR-1248-4

R.T.: 6.590 min
Delta R.T.: -0.002 min
Response: 8120629
Conc: 104.99 ng/ml



#24 AR-1248-4


R.T.: 5.223 min
Delta R.T.: -0.005 min
Response: 22495313
Conc: 334.93 ng/ml



#25 AR-1248-5


R.T.: 6.630 min
Delta R.T.: -0.001 min
Response: 7155447
Conc: 95.10 ng/ml

Instrument :
FID_C
ClientSampleId :




#25 AR-1248-5

R.T.: 5.620 min
Delta R.T.: -0.006 min
Response: 6170905
Conc: 93.55 ng/ml




#26 AR-1254-1

R.T.: 6.564 min
Delta R.T.: -0.001 min
Response: 21317266
Conc: 267.01 ng/ml



#26 AR-1254-1


R.T.: 5.579 min
Delta R.T.: -0.006 min
Response: 22660193
Conc: 225.80 ng/ml



#27 AR-1254-2


R.T.: 6.786 min
Delta R.T.: -0.001 min
Response: 18233954
Conc: 152.24 ng/ml

Instrument :
FID_C
ClientSampleId :




#27 AR-1254-2

R.T.: 5.728 min
Delta R.T.: -0.006 min
Response: 17684677
Conc: 203.56 ng/ml




#28 AR-1254-3

R.T.: 7.156 min
Delta R.T.: -0.002 min
Response: 9011770
Conc: 74.18 ng/ml



#28 AR-1254-3


R.T.: 6.152 min
Delta R.T.: 0.010 min
Response: 30509221
Conc: 217.30 ng/ml



#29 AR-1254-4


R.T.: 7.445 min
Delta R.T.: -0.001 min
Response: 5934839
Conc: 66.74 ng/ml

Instrument :
FID_C
ClientSampleId :




#29 AR-1254-4

R.T.: 6.366 min
Delta R.T.: -0.006 min
Response: 2294131
Conc: 27.25 ng/ml




#30 AR-1254-5

R.T.: 7.867 min
Delta R.T.: -0.004 min
Response: 56362780
Conc: 579.42 ng/ml



#30 AR-1254-5


R.T.: 6.787 min
Delta R.T.: -0.009 min
Response: 53760290
Conc: 453.22 ng/ml



#31 AR-1260-1


R.T.: 7.323 min
Delta R.T.: -0.003 min
Response: 44062388
Conc: 488.83 ng/ml

Instrument :
FID_C
ClientSampleId :




#31 AR-1260-1

R.T.: 6.267 min
Delta R.T.: -0.010 min
Response: 44116612
Conc: 473.78 ng/ml




#32 AR-1260-2

R.T.: 7.582 min
Delta R.T.: -0.003 min
Response: 49514400
Conc: 482.65 ng/ml



#32 AR-1260-2


R.T.: 6.456 min
Delta R.T.: -0.010 min
Response: 51075141
Conc: 465.86 ng/ml



#33 AR-1260-3


R.T.: 7.945 min
Delta R.T.: -0.003 min
Response: 38481553
Conc: 488.60 ng/ml

Instrument :
FID_C
ClientSampleId :




#33 AR-1260-3

R.T.: 6.612 min
Delta R.T.: -0.010 min
Response: 46377440
Conc: 450.90 ng/ml




#34 AR-1260-4

R.T.: 8.175 min
Delta R.T.: -0.003 min
Response: 43726248
Conc: 487.91 ng/ml



#34 AR-1260-4


R.T.: 7.088 min
Delta R.T.: -0.011 min
Response: 39317813
Conc: 472.09 ng/ml



#35 AR-1260-5


R.T.: 8.509 min
Delta R.T.: -0.004 min
Response: 79499916
Conc: 465.84 ng/ml

Instrument :
FID_C
ClientSampleId :




#35 AR-1260-5

R.T.: 7.330 min
Delta R.T.: -0.011 min
Response: 87024460
Conc: 452.30 ng/ml




#36 AR-1262-1

R.T.: 7.945 min
Delta R.T.: -0.002 min
Response: 38481553
Conc: 332.50 ng/ml



#36 AR-1262-1


R.T.: 6.880 min
Delta R.T.: -0.008 min
Response: 20955776
Conc: 415.18 ng/ml



#37 AR-1262-2


R.T.: 8.509 min
Delta R.T.: -0.002 min
Response: 79499916
Conc: 397.12 ng/ml

Instrument :
FID_C
ClientSampleId :




#37 AR-1262-2

R.T.: 7.088 min
Delta R.T.: -0.008 min
Response: 39317813
Conc: 348.02 ng/ml




#38 AR-1262-3

R.T.: 8.849 min
Delta R.T.: 0.009 min
Response: 53608888
Conc: 372.28 ng/ml



#38 AR-1262-3


R.T.: 7.618 min
Delta R.T.: -0.008 min
Response: 19294845
Conc: 220.12 ng/ml



#39 AR-1262-4


R.T.: 8.912 min
Delta R.T.: -0.021 min
Response: 36049493
Conc: 317.46 ng/ml

Instrument :
FID_C
ClientSampleId :




#39 AR-1262-4

R.T.: 7.681 min
Delta R.T.: -0.009 min
Response: 60590596
Conc: 371.02 ng/ml




#40 AR-1262-5

R.T.: 9.624 min
Delta R.T.: -0.001 min
Response: 23023779
Conc: 308.62 ng/ml



#40 AR-1262-5


R.T.: 8.181 min
Delta R.T.: -0.009 min
Response: 19479965
Conc: 260.97 ng/ml



#41 AR-1268-1


R.T.: 8.849 min
Delta R.T.: 0.011 min
Response: 53608888
Conc: 213.79 ng/ml

Instrument :
FID_C
ClientSampleId :




#41 AR-1268-1

R.T.: 7.618 min
Delta R.T.: -0.008 min
Response: 19294845
Conc: 74.11 ng/ml




#42 AR-1268-2

R.T.: 8.912 min
Delta R.T.: -0.024 min
Response: 36049493
Conc: 157.08 ng/ml



#42 AR-1268-2


R.T.: 7.681 min
Delta R.T.: -0.009 min
Response: 60590596
Conc: 262.08 ng/ml



#43 AR-1268-3


R.T.: 9.175 min
Delta R.T.: -0.004 min
Response: 1160014
Conc: 5.88 ng/ml

Instrument :
FID_C
ClientSampleId :




#43 AR-1268-3

R.T.: 7.893 min
Delta R.T.: -0.008 min
Response: 1100834
Conc: 5.39 ng/ml




#44 AR-1268-4

R.T.: 9.624 min
Delta R.T.: -0.002 min
Response: 23023779
Conc: 277.97 ng/ml



#44 AR-1268-4


R.T.: 8.181 min
Delta R.T.: -0.008 min
Response: 19479965
Conc: 241.25 ng/ml



#45 AR-1268-5

R.T.: 10.065 min
Delta R.T.: -0.004 min
Response: 6443604
Conc: 10.65 ng/ml

Instrument :
FID_C
ClientSampleId :



#45 AR-1268-5

R.T.: 8.479 min
Delta R.T.: -0.010 min
Response: 6014460
Conc: 9.52 ng/ml