

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC092024AL\
 Data File : FC067199.D
 Signal(s) : FID1A.ch
 Acq On : 20 Sep 2024 11:19
 Operator : YP/AJ
 Sample : P4104-06DL 10X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 WC-14-5-7.5B-EPH-1DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/23/2024
 Supervised By :Ankita Jodhani 09/23/2024

Integration File: autoint1.e
 Quant Time: Sep 21 03:18:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 091024.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 11 02:55:44 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.940	241164	2.469 ug/mlm
Spiked Amount	50.000	Recovery =	4.94%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC092024AL\
Data File : FC067199.D
Signal(s) : FID1A.ch
Acq On : 20 Sep 2024 11:19
Operator : YP/AJ
Sample : P4104-06DL 10X
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
WC-14-5-7.5B-EPH-1DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/23/2024
Supervised By :Ankita Jodhani 09/23/2024

Integration File: autoint1.e
Quant Time: Sep 21 03:18:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 091024.M
Quant Title : GC Extractables
QLast Update : Wed Sep 11 02:55:44 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.504 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :
WC-14-5-7.5B-EPH-1DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/23/2024
Supervised By :Ankita Jodhani 09/23/2024



#12 1-chlorooctadecane (SURR)

R.T.: 12.940 min
Delta R.T.: -0.003 min
Response: 241164
Conc: 2.47 ug/ml m