

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC092120AL\
 Data File : FC049676.D
 Signal(s) : FID1A.CH
 Acq On : 22 Sep 2020 1:34
 Operator : DD\AJ
 Sample : L3869-08DL 5X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Sep 22 03:50:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 090820.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 09 05:58:47 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.661	606346	5.277 ug/ml
Spiked Amount 50.000		Recovery =	10.55%
12) S 1-chlorooctadecane (S...	13.094	544431	6.112 ug/ml
Spiked Amount 50.000		Recovery =	12.22%
Target Compounds			
10) T n-Eicosane (C20)	12.721f	360586	3.367 ug/ml
14) T n-Tetracosane (C24)	14.939	158498	1.515 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC092120AL\
Data File : FC049676.D
Signal(s) : FID1A.CH
Acq On : 22 Sep 2020 1:34
Operator : DD\AJ
Sample : L3869-08DL 5X
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Sep 22 03:50:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 090820.M
Quant Title : GC Extractables
QLast Update : Wed Sep 09 05:58:47 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.661 min
Delta R.T.: -0.018 min
Response: 606346
Conc: 5.28 ug/ml

Instrument :
FID_C
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.721 min
Delta R.T.: 0.057 min
Response: 360586
Conc: 3.37 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 13.094 min
Delta R.T.: -0.015 min
Response: 544431
Conc: 6.11 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.939 min
Delta R.T.: -0.025 min
Response: 158498
Conc: 1.51 ug/ml