

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC09292
Data File : FC069866.D
Signal(s) : FID1A.ch
Acq On : 29 Sep 2025 16:06
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Reviewed By :Yogesh Patel 09/30/2025
Supervised By :mohammad ahmed 10/01/2025

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 092325.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.388	3.334	3.445	BB	313566	2590471	71.84%	3.978%
2	4.459	4.399	4.500	BB	308995	2766612	76.72%	4.248%
3	6.052	6.000	6.127	BB	333746	3100726	85.99%	4.761%
4	6.481	6.427	6.527	BB	299397	2978681	82.60%	4.574%
5	7.110	7.062	7.212	BB	317177	3106401	86.14%	4.770%
6	8.277	8.222	8.332	BB	296147	3113781	86.35%	4.781%
7	9.881	9.824	9.952	BB	307039	3501385	97.10%	5.377%
8	11.325	11.264	11.380	BB	296165	3442637	95.47%	5.286%
9	11.648	11.589	11.689	BB	333880	3606082	100.00%	5.537%
10	12.636	12.569	12.692	BB	291343	3330453	92.36%	5.114%
11	13.082	13.009	13.125	BB	237261	2877860	79.81%	4.419%
12	13.248	13.210	13.302	BB	270735	3269850	90.68%	5.021%
13	13.834	13.799	13.889	VB	261504	3305672	91.67%	5.076%
14	14.938	14.874	14.995	BB	250024	3236678	89.76%	4.970%
15	15.960	15.890	16.010	BB	243568	3142406	87.14%	4.825%
16	16.913	16.830	16.948	BB	227547	3080061	85.41%	4.730%
17	17.800	17.709	17.837	VV	235950	3190660	88.48%	4.899%
18	18.634	18.589	18.669	M	228788	3092570	85.76%	4.749%
19	19.416	19.367	19.463	M	184944	2706870	75.06%	4.157%
20	20.157	20.121	20.204	BV	154138	2198220	60.96%	3.376%
21	20.913	20.881	20.985	BV	100901	1780450	49.37%	2.734%
22	21.866	21.804	21.926	PV	68021	1703885	47.25%	2.616%
Sum of corrected areas:						65122411		

Aliphatic EPH 092325.M Tue Sep 30 06:14:18 2025