

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 10/02/2025
Supervised By :mohammad ahmed 10/04/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC10022
Data File : FC069894.D
Signal(s) : FID1A.ch
Acq On : 01 Oct 2025 10:24
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 092325.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.385	3.349	3.432	BB	280550	2347334	65.57%	3.678%
2	4.456	4.420	4.519	BB	303754	2715993	75.87%	4.256%
3	6.048	6.010	6.134	BB	324426	3061173	85.51%	4.797%
4	6.478	6.445	6.525	BB	303297	2941246	82.16%	4.609%
5	7.107	7.065	7.219	BB	316174	3089165	86.29%	4.841%
6	8.275	8.219	8.352	BB	294574	3110537	86.89%	4.874%
7	9.878	9.819	9.959	BB	306109	3361161	93.89%	5.267%
8	11.322	11.204	11.392	BB	303404	3461722	96.70%	5.424%
9	11.643	11.582	11.705	BB	337232	3579984	100.00%	5.610%
10	12.633	12.565	12.684	BB	291118	3296235	92.07%	5.165%
11	13.078	13.017	13.137	BB	238512	2809520	78.48%	4.402%
12	13.245	13.177	13.292	BB	270065	3203601	89.49%	5.020%
13	13.832	13.799	13.872	VB	257373	3195454	89.26%	5.007%
14	14.936	14.902	14.992	VB	258633	3182958	88.91%	4.988%
15	15.958	15.892	16.017	BB	229970	3076973	85.95%	4.821%
16	16.908	16.830	16.984	BB	227041	3062914	85.56%	4.799%
17	17.796	17.724	17.900	BB	229580	3309251	92.44%	5.185%
18	18.630	18.581	18.720	M	209228	3285426	91.77%	5.148%
19	19.412	19.359	19.537	BV	176657	2775332	77.52%	4.349%
20	20.152	20.120	20.255	BV	126881	2018775	56.39%	3.163%
21	20.909	20.863	20.973	M	83670	1568640	43.82%	2.458%
22	21.860	21.772	21.984	BB	52545	1364593	38.12%	2.138%
Sum of corrected areas:							63817987	

Aliphatic EPH 092325.M Wed Oct 01 16:37:46 2025