

Instrument :
FID_C
LabSampleId :
10 PPM ALIPHATIC HC STD4

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 10/11/2023
Supervised By :Ankita Jodhani 10/11/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\F0
Data File : FC064786.D
Signal(s) : FID1A.ch
Acq On : 10 Oct 2023 11:03
Sample : 10 PPM ALIPHATIC HC STD4
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101023.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.321	3.283	3.416	BB	215245	1995053	86.68%	5.154%
2	4.381	4.341	4.481	BB	202474	1976877	85.89%	5.107%
3	5.968	5.931	6.131	BB	121894	1948379	84.65%	5.033%
4	6.391	6.348	6.494	BB	187531	2027480	88.09%	5.237%
5	7.022	6.927	7.191	BB	96848	1738416	75.53%	4.491%
6	8.179	8.136	8.286	BB	154274	1936495	84.13%	5.002%
7	9.773	9.727	9.875	BB	151198	1914901	83.19%	4.947%
8	11.208	11.176	11.295	BB	148616	1941562	84.35%	5.015%
9	11.522	11.476	11.616	BB	195714	2301718	100.00%	5.946%
10	12.511	12.462	12.580	BB	154885	1864335	81.00%	4.816%
11	12.954	12.911	13.043	BB	123115	1611806	70.03%	4.164%
12	13.120	13.069	13.196	BB	141943	1833204	79.65%	4.736%
13	13.703	13.655	13.782	BB	132341	1775292	77.13%	4.586%
14	14.802	14.751	14.886	BB	123633	1713867	74.46%	4.427%
15	15.818	15.764	15.896	BB	118654	1646936	71.55%	4.254%
16	16.763	16.712	16.831	BB	114797	1582940	68.77%	4.089%
17	17.646	17.591	17.715	BB	107330	1571459	68.27%	4.059%
18	18.477	18.425	18.535	BB	103967	1554029	67.52%	4.014%
19	19.255	19.198	19.330	BB	101181	1518856	65.99%	3.924%
20	19.991	19.778	20.067	BB	94329	1507714	65.50%	3.895%
21	20.714	20.650	20.797	BB	74517	1406500	61.11%	3.633%
22	21.595	21.530	21.743	BB	52616	1343630	58.38%	3.471%
Sum of corrected areas:						38711449		

Aliphatic EPH 101023.M Tue Oct 10 15:07:04 2023