

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC101119AL\
 Data File : FC043917.D
 Signal(s) : FID1A.CH
 Acq On : 11 Oct 2019 21:45
 Operator : DD\AJ
 Sample : K5316-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Oct 12 00:07:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 092519.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 12 00:05:39 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.569	5275881	34.760 ug/ml
Spiked Amount 50.000		Recovery =	69.52%
12) S 1-chlorooctadecane (S...	12.988	4054099	34.194 ug/ml
Spiked Amount 50.000		Recovery =	68.39%
Target Compounds			
10) T n-Eicosane (C20)	12.517	249722	1.789 ug/ml
13) T n-Docosane (C22)	13.674f	251252	1.768 ug/ml
14) T n-Tetracosane (C24)	14.845	1903245	13.258 ug/ml
16) T n-Octacosane (C28)	16.750	1655684	11.680 ug/ml
20) T n-Hexatriacontane (C36)	19.949f	231417	2.002 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC101119AL\
Data File : FC043917.D
Signal(s) : FID1A.CH
Acq On : 11 Oct 2019 21:45
Operator : DD\AJ
Sample : K5316-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Oct 12 00:07:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 092519.M
Quant Title : GC Extractables
QLast Update : Sat Oct 12 00:05:39 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.569 min
Delta R.T.: 0.002 min
Response: 5275881
Conc: 34.76 ug/ml

Instrument :
FID_C
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.517 min
Delta R.T.: -0.027 min
Response: 249722
Conc: 1.79 ug/ml




#12 1-chlorooctadecane (SURR)

R.T.: 12.988 min
Delta R.T.: 0.002 min
Response: 4054099
Conc: 34.19 ug/ml



#13 n-Docosane (C22)


R.T.: 13.674 min
Delta R.T.: -0.057 min
Response: 251252
Conc: 1.77 ug/ml



#14 n-Tetracosane (C24)


R.T.: 14.845 min
Delta R.T.: 0.019 min
Response: 1903245
Conc: 13.26 ug/ml

Instrument :
FID_C
ClientSampleId :



#16 n-Octacosane (C28)

R.T.: 16.750 min
Delta R.T.: -0.034 min
Response: 1655684
Conc: 11.68 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 19.949 min
Delta R.T.: -0.058 min
Response: 231417
Conc: 2.00 ug/ml