

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC101221AL\  
 Data File : FC056141.D  
 Signal(s) : FID1A.CH  
 Acq On : 12 Oct 2021 22:35  
 Operator : AJ\MA  
 Sample : M4163-01DL 5X  
 Misc :  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 FID\_C  
 ClientSampleId :  
 TR-04-100821DL

Integration File: sample.E  
 Quant Time: Oct 13 05:08:21 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 092921.M  
 Quant Title : GC Extractables  
 QLast Update : Thu Sep 30 01:25:44 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response | Conc Units    |
|--------------------------------|--------|----------|---------------|
| -----                          |        |          |               |
| System Monitoring Compounds    |        |          |               |
| 9) S ortho-Terphenyl (SURR)    | 11.636 | 795637   | 6.705 ug/ml   |
| Spiked Amount                  | 50.000 | Recovery | = 13.41%      |
| 12) S 1-chlorooctadecane (S... | 13.069 | 20857645 | 227.027 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 454.05%     |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC101221AL\  
Data File : FC056141.D  
Signal(s) : FID1A.CH  
Acq On : 12 Oct 2021 22:35  
Operator : AJ\MA  
Sample : M4163-01DL 5X  
Misc :  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :  
TR-04-100821DL

Integration File: sample.E  
Quant Time: Oct 13 05:08:21 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 092921.M  
Quant Title : GC Extractables  
QLast Update : Thu Sep 30 01:25:44 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.636 min  
Delta R.T.: -0.026 min  
Response: 795637  
Conc: 6.71 ug/ml

Instrument :  
FID\_C  
ClientSampleId :  
TR-04-100821DL



#12 1-chlorooctadecane (SURR)

R.T.: 13.069 min  
Delta R.T.: -0.023 min  
Response: 20857645  
Conc: 227.03 ug/ml