

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC101323AL\
Data File : FC064820.D
Signal(s) : FID1A.ch
Acq On : 14 Oct 2023 01:33
Operator : YP/AJ
Sample : 04816-03
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
T-3

Integration File: autoint1.e
Quant Time: Oct 14 03:01:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101023.M
Quant Title : GC Extractables
QLast Update : Tue Oct 10 12:17:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.953	7653434	51.043 ug/ml
Spiked Amount 50.000		Recovery	= 102.09%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC101323AL\
Data File : FC064820.D
Signal(s) : FID1A.ch
Acq On : 14 Oct 2023 01:33
Operator : YP/AJ
Sample : 04816-03
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
T-3

Integration File: autoint1.e
Quant Time: Oct 14 03:01:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101023.M
Quant Title : GC Extractables
QLast Update : Tue Oct 10 12:17:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.521 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :
T-3



#12 1-chlorooctadecane (SURR)

R.T.: 12.953 min
Delta R.T.: -0.001 min
Response: 7653434
Conc: 51.04 ug/ml