

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC102121AL\
 Data File : FC056298.D
 Signal(s) : FID1A.CH
 Acq On : 21 Oct 2021 00:47
 Operator : AJ\MA
 Sample : 20 PPM ALIPHATIC HC STD ICV
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC HC STD ICV

Integration File: sample.E
 Quant Time: Oct 21 01:28:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102121.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 21 01:00:25 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	108.234	108.213 E3	0.0	100	0.00
2 T n-Decane (C10)	107.719	108.367 E3	-0.6	100	0.00
3 T A~Naphthalene (C11.7)	121.921	122.009 E3	-0.1	100	0.00
4 T n-Dodecane (C12)	110.239	110.213 E3	0.0	100	0.00
5 T A~2-methylnaphthalene (C12.	122.461	123.453 E3	-0.8	101	0.00
6 T n-Tetradecane (C14)	112.268	111.949 E3	0.3	100	0.00
7 T n-Hexadecane (C16)	113.690	113.767 E3	-0.1	101	0.00
8 T n-Octadecane (C18)	116.858	117.379 E3	-0.4	101	0.00
9 S ortho-Terphenyl (SURR)	123.818	124.167 E3	-0.3	101	0.00
10 T n-Eicosane (C20)	114.927	115.386 E3	-0.4	101	0.00
11 T n-Heneicosane (C21)	114.923	115.401 E3	-0.4	100	0.00
12 S 1-chlorooctadecane (SURR)	96.224	96.706 E3	-0.5	100	0.00
13 T n-Docosane (C22)	112.585	113.754 E3	-1.0	101	0.00
14 T n-Tetracosane (C24)	110.204	111.583 E3	-1.3	101	0.00
15 T n-Hexacosane (C26)	105.545	106.863 E3	-1.2	101	0.00
16 T n-Octacosane (C28)	103.727	103.623 E3	0.1	100	0.00
17 T n-Tricontane (C30)	103.608	100.569 E3	2.9	99	0.00
18 T n-Dotriacontane (C32)	95.809	93.648 E3	2.3	99	0.00
19 T n-Tetratriacontane (C34)	87.709	88.026 E3	-0.4	101	0.00
20 T n-Hexatriacontane (C36)	81.852	82.082 E3	-0.3	101	0.00
21 T n-Octatriacontane (C38)	79.976	80.158 E3	-0.2	101	0.00
22 T n-Tetracontane (C40)	82.390	81.280 E3	1.3	99	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0