

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC102324AL\
 Data File : FC067508.D
 Signal(s) : FID1A.ch
 Acq On : 23 Oct 2024 16:25
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD3
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
 Quant Time: Oct 23 17:17:15 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102324.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 23 17:16:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.496	2885453	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
12) S 1-chlorooctadecane (S...	12.935	2131037	20.000 ug/ml
Spiked Amount 50.000		Recovery =	40.00%
Target Compounds			
1) T n-Nonane (C9)	3.275	2673666	20.000 ug/ml
2) T n-Decane (C10)	4.339	2702691	20.000 ug/ml
3) T A~Naphthalene (C11.7)	5.916	2941492	20.000 ug/ml
4) T n-Dodecane (C12)	6.354	2685213	20.000 ug/ml
5) T A~2-methylnaphthalene...	6.971	2804469	20.000 ug/ml
6) T n-Tetradecane (C14)	8.146	2596600	20.000 ug/ml
7) T n-Hexadecane (C16)	9.745	2611645	20.000 ug/ml
8) T n-Octadecane (C18)	11.185	2610385	20.000 ug/ml
10) T n-Eicosane (C20)	12.492	2486010	20.000 ug/ml
11) T n-Heneicosane (C21)	13.103	2429615	20.000 ug/ml
13) T n-Docosane (C22)	13.688	2410406	20.000 ug/ml
14) T n-Tetracosane (C24)	14.790	2385181	20.000 ug/ml
15) T n-Hexacosane (C26)	15.809	2344463	20.000 ug/ml
16) T n-Octacosane (C28)	16.757	2318777	20.000 ug/ml
17) T n-Tricontane (C30)	17.644	2378909	20.000 ug/ml
18) T n-Dotriacontane (C32)	18.475	2323289	20.000 ug/ml
19) T n-Tetratriacontane (C34)	19.258	2065430	20.000 ug/ml
20) T n-Hexatriacontane (C36)	19.995	1794670	20.000 ug/ml
21) T n-Octatriacontane (C38)	20.718	1712685	20.000 ug/ml
22) T n-Tetracontane (C40)	21.604	1679680	20.000 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC102324AL\
Data File : FC067508.D
Signal(s) : FID1A.ch
Acq On : 23 Oct 2024 16:25
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD3
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD3

Integration File: autoint1.e
Quant Time: Oct 23 17:17:15 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102324.M
Quant Title : GC Extractables
QLast Update : Wed Oct 23 17:16:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

