

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC102722AL\
 Data File : FC061977.D
 Signal(s) : FID1A.ch
 Acq On : 28 Oct 2022 01:41
 Operator : YP/AJ
 Sample : N5273-01
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 LIDEN-GAS-PIPELINE

Integration File: autoint1.e
 Quant Time: Oct 28 06:15:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101722.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 17 23:26:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.591	6070075	40.837 ug/ml
Spiked Amount 50.000		Recovery =	81.67%
12) S 1-chlorooctadecane (S...	13.027	4841753	41.560 ug/ml
Spiked Amount 50.000		Recovery =	83.12%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC102722AL\
Data File : FC061977.D
Signal(s) : FID1A.ch
Acq On : 28 Oct 2022 01:41
Operator : YP/AJ
Sample : N5273-01
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
LIDEN-GAS-PIPELINE

Integration File: autoint1.e
Quant Time: Oct 28 06:15:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101722.M
Quant Title : GC Extractables
QLast Update : Mon Oct 17 23:26:31 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.591 min
Delta R.T.: 0.006 min
Response: 6070075
Conc: 40.84 ug/ml

Instrument :
FID_C
ClientSampleId :
LIDEN-GAS-PIPELINE



#12 1-chlorooctadecane (SURR)

R.T.: 13.027 min
Delta R.T.: 0.003 min
Response: 4841753
Conc: 41.56 ug/ml