

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC102822AL\
 Data File : FC061996.D
 Signal(s) : FID1A.ch
 Acq On : 28 Oct 2022 15:56
 Operator : YP/AJ
 Sample : N5267-18DL 20X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 TP6-1024-EPH-2DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/29/2022
 Supervised By :Ankita Jodhani 10/31/2022

Integration File: autoint1.e
 Quant Time: Oct 31 11:32:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101722.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 17 23:26:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.583	266924	1.796 ug/mlm
Spiked Amount 50.000		Recovery =	3.59%
12) S 1-chlorooctadecane (S...	13.019	248062	2.129 ug/mlm
Spiked Amount 50.000		Recovery =	4.26%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC102822AL\
Data File : FC061996.D
Signal(s) : FID1A.ch
Acq On : 28 Oct 2022 15:56
Operator : YP/AJ
Sample : N5267-18DL 20X
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
TP6-1024-EPH-2DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2022
Supervised By :Ankita Jodhani 10/31/2022

Integration File: autoint1.e
Quant Time: Oct 31 11:32:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101722.M
Quant Title : GC Extractables
QLast Update : Mon Oct 17 23:26:31 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.583 min
Delta R.T.: -0.003 min
Response: 266924
Conc: 1.80 ug/ml

Instrument :
FID_C
ClientSampleId :
TP6-1024-EPH-2DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/29/2022
Supervised By :Ankita Jodhani 10/31/2022



#12 1-chlorooctadecane (SURR)

R.T.: 13.019 min
Delta R.T.: -0.004 min
Response: 248062
Conc: 2.13 ug/ml m