

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC103024AL\
 Data File : FC067588.D
 Signal(s) : FID1A.ch
 Acq On : 31 Oct 2024 07:23
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Nov 01 00:22:40 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102324.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 23 18:12:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.493	2817637	19.203 ug/ml
Spiked Amount 50.000		Recovery =	38.41%
12) S 1-chlorooctadecane (S...	12.930	2106666	19.508 ug/ml
Spiked Amount 50.000		Recovery =	39.02%
Target Compounds			
1) T n-Nonane (C9)	3.274	2562501	19.021 ug/ml
2) T n-Decane (C10)	4.338	2611307	19.166 ug/ml
3) T A~Naphthalene (C11.7)	5.915	2826595	19.074 ug/ml
4) T n-Dodecane (C12)	6.353	2604642	19.231 ug/ml
5) T A~2-methylnaphthalene...	6.969	2695981	19.090 ug/ml
6) T n-Tetradecane (C14)	8.144	2532295	19.292 ug/ml
7) T n-Hexadecane (C16)	9.742	2562151	19.395 ug/ml
8) T n-Octadecane (C18)	11.181	2579000	19.521 ug/ml
10) T n-Eicosane (C20)	12.489	2455785	19.463 ug/ml
11) T n-Heneicosane (C21)	13.099	2405500	19.491 ug/ml
13) T n-Docosane (C22)	13.684	2388418	19.505 ug/ml
14) T n-Tetracosane (C24)	14.784	2378410	19.622 ug/ml
15) T n-Hexacosane (C26)	15.803	2349366	19.733 ug/ml
16) T n-Octacosane (C28)	16.752	2380755	20.236 ug/ml
17) T n-Tricontane (C30)	17.638	2551686	21.111 ug/ml
18) T n-Dotriacontane (C32)	18.469	2563153	21.663 ug/ml
19) T n-Tetratriacontane (C34)	19.250	2255385	21.536 ug/ml
20) T n-Hexatriacontane (C36)	19.988	1916156	21.050 ug/ml
21) T n-Octatriacontane (C38)	20.711	1797274	20.753 ug/ml
22) T n-Tetracontane (C40)	21.594	1737866	20.443 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC103024AL\
Data File : FC067588.D
Signal(s) : FID1A.ch
Acq On : 31 Oct 2024 07:23
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
Quant Time: Nov 01 00:22:40 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102324.M
Quant Title : GC Extractables
QLast Update : Wed Oct 23 18:12:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

