

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110222AL\
 Data File : FC062066.D
 Signal(s) : FID1A.ch
 Acq On : 02 Nov 2022 10:50
 Operator : YP/AJ
 Sample : N5392-02DL 10X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 OE-1-103122-EPH-2DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/03/2022
 Supervised By :Ankita Jodhani 11/03/2022

Integration File: autoint1.e
 Quant Time: Nov 03 06:17:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101722.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 17 23:26:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.582	363830	2.448 ug/mlm
Spiked Amount 50.000		Recovery =	4.90%
12) S 1-chlorooctadecane (S...	13.006	777725	6.676 ug/mlm
Spiked Amount 50.000		Recovery =	13.35%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110222AL\
Data File : FC062066.D
Signal(s) : FID1A.ch
Acq On : 02 Nov 2022 10:50
Operator : YP/AJ
Sample : N5392-02DL 10X
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
OE-1-103122-EPH-2DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/03/2022
Supervised By :Ankita Jodhani 11/03/2022

Integration File: autoint1.e
Quant Time: Nov 03 06:17:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101722.M
Quant Title : GC Extractables
QLast Update : Mon Oct 17 23:26:31 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.582 min
Delta R.T.: -0.003 min
Response: 363830
Conc: 2.45 ug/ml

Instrument :
FID_C
ClientSampleId :
OE-1-103122-EPH-2DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/03/2022
Supervised By :Ankita Jodhani 11/03/2022



#12 1-chlorooctadecane (SURR)

R.T.: 13.006 min
Delta R.T.: -0.017 min
Response: 777725
Conc: 6.68 ug/ml m