

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110723AL\
 Data File : FC065054.D
 Signal(s) : FID1A.ch
 Acq On : 08 Nov 2023 06:37
 Operator : YP/AJ
 Sample : 05251-11
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Nov 08 07:53:46 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 28 02:12:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.946	6993527	44.239 ug/ml
Spiked Amount 50.000		Recovery =	88.48%
Target Compounds			
4) T n-Dodecane (C12)	6.382	36395	0.188 ug/ml
5) T A~2-methylnaphthalene...	6.997	48840	0.266 ug/ml
17) T n-Tricontane (C30)	17.637	53414	0.337 ug/ml
18) T n-Dotriacontane (C32)	18.465	48381	0.306 ug/ml
19) T n-Tetratriacontane (C34)	19.247	47919	0.314 ug/ml
20) T n-Hexatriacontane (C36)	19.982	46238	0.306 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110723AL\
Data File : FC065054.D
Signal(s) : FID1A.ch
Acq On : 08 Nov 2023 06:37
Operator : YP/AJ
Sample : 05251-11
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Nov 08 07:53:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
Quant Title : GC Extractables
QLast Update : Sat Oct 28 02:12:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#4 n-Dodecane (C12)

R.T.: 6.382 min
Delta R.T.: -0.010 min
Response: 36395
Conc: 0.19 ug/ml

Instrument :


FID_C

ClientSampleId :




#5 A~2-methylnaphthalene (C12.89)

R.T.: 6.997 min
Delta R.T.: -0.019 min
Response: 48840
Conc: 0.27 ug/ml




#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.523 min
Response: 0
Conc: N.D.



#12 1-chlorooctadecane (SURR)


R.T.: 12.946 min
Delta R.T.: -0.009 min
Response: 6993527
Conc: 44.24 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.637 min
Delta R.T.: -0.012 min
Response: 53414
Conc: 0.34 ug/ml

Instrument :
FID_C
ClientSampleId :




#18 n-Dotriacontane (C32)

R.T.: 18.465 min
Delta R.T.: -0.012 min
Response: 48381
Conc: 0.31 ug/ml



#19 n-Tetatriacontane (C34)

R.T.: 19.247 min
Delta R.T.: -0.009 min
Response: 47919
Conc: 0.31 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 19.982 min
Delta R.T.: -0.012 min
Response: 46238
Conc: 0.31 ug/ml