

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110723AL\
 Data File : FC065049.D
 Signal(s) : FID1A.ch
 Acq On : 08 Nov 2023 03:31
 Operator : YP/AJ
 Sample : 05251-06
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 T-6

Integration File: autoint1.e
 Quant Time: Nov 08 05:09:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 28 02:12:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.947	8504882	53.800 ug/ml
Spiked Amount	50.000	Recovery	= 107.60%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110723AL\
Data File : FC065049.D
Signal(s) : FID1A.ch
Acq On : 08 Nov 2023 03:31
Operator : YP/AJ
Sample : 05251-06
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
T-6

Integration File: autoint1.e
Quant Time: Nov 08 05:09:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
Quant Title : GC Extractables
QLast Update : Sat Oct 28 02:12:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um





#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.523 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :
T-6

#12 1-chlorooctadecane (SURR)

R.T.: 12.947 min
Delta R.T.: -0.008 min
Response: 8504882
Conc: 53.80 ug/ml