

nteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC11072
Data File : FC067666.D
Signal(s) : FID1A.ch
Acq On : 07 Nov 2024 22:09
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/08/2024
Supervised By :Ankita Jodhani 11/08/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102324.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.271	3.245	3.319	BB	325340	2763672	90.42%	4.987%
2	4.335	4.274	4.404	BB	304167	2812074	92.01%	5.074%
3	5.913	5.857	5.989	BB	332453	3056372	100.00%	5.515%
4	6.350	6.285	6.409	BB	279476	2801824	91.67%	5.055%
5	6.968	6.920	7.072	BB	289872	2927765	95.79%	5.283%
6	8.142	8.077	8.207	BB	258630	2728142	89.26%	4.922%
7	9.740	9.682	9.802	BB	245619	2752720	90.06%	4.967%
8	11.180	11.062	11.235	BB	231651	2743032	89.75%	4.949%
9	11.491	11.435	11.554	BB	274911	2974319	97.32%	5.367%
10	12.487	12.419	12.535	BB	216455	2600353	85.08%	4.692%
11	12.929	12.864	12.977	BB	185867	2234180	73.10%	4.031%
12	13.098	13.035	13.145	BB	200578	2545253	83.28%	4.592%
13	13.683	13.614	13.745	BB	200651	2523166	82.55%	4.553%
14	14.784	14.717	14.844	BB	192945	2497861	81.73%	4.507%
15	15.804	15.740	15.865	BB	181344	2455745	80.35%	4.431%
16	16.752	16.674	16.815	BB	174112	2424021	79.31%	4.374%
17	17.637	17.564	17.684	BB	178588	2475521	81.00%	4.467%
18	18.469	18.402	18.529	BB	170482	2415091	79.02%	4.358%
19	19.251	19.180	19.322	BB	150050	2167698	70.92%	3.911%
20	19.990	19.714	20.055	BB	126675	1943255	63.58%	3.506%
21	20.712	20.630	20.785	BB	102641	1822443	59.63%	3.288%
22	21.594	21.502	21.674	BB	72587	1757565	57.50%	3.171%
Sum of corrected areas:						55422069		

Aliphatic EPH 102324.M Fri Nov 08 01:24:19 2024