

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110819AL\
 Data File : FC044430.D
 Signal(s) : FID1A.CH
 Acq On : 09 Nov 2019 00:02
 Operator : DD\AJ
 Sample : K5730-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 HILTON-STOCKPILE

Integration File: autoint1.e
 Quant Time: Nov 09 02:14:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 102219.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 22 12:36:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.720	4814490	33.030 ug/ml
Spiked Amount 50.000		Recovery =	66.06%
12) S 1-chlorooctadecane (S...	13.150	4104420	36.708 ug/ml
Spiked Amount 50.000		Recovery =	73.42%
Target Compounds			
14) T n-Tetracosane (C24)	14.993	200643	1.774 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110819AL\
Data File : FC044430.D
Signal(s) : FID1A.CH
Acq On : 09 Nov 2019 00:02
Operator : DD\AJ
Sample : K5730-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
HILTON-STOCKPILE

Integration File: autoint1.e
Quant Time: Nov 09 02:14:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 102219.M
Quant Title : GC Extractables
QLast Update : Tue Oct 22 12:36:58 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.720 min
Delta R.T.: 0.012 min
Response: 4814490
Conc: 33.03 ug/ml

Instrument :
FID_C
ClientSampleId :
HILTON-STOCKPILE



#12 1-chlorooctadecane (SURR)

R.T.: 13.150 min
Delta R.T.: 0.008 min
Response: 4104420
Conc: 36.71 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.993 min
Delta R.T.: -0.008 min
Response: 200643
Conc: 1.77 ug/ml