

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110822AL\
 Data File : FC062111.D
 Signal(s) : FID1A.ch
 Acq On : 08 Nov 2022 18:30
 Operator : YP/AJ
 Sample : N5506-13
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 TP-10

Integration File: autoint1.e
 Quant Time: Nov 09 03:47:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101722.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 17 23:26:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.594	4334922	29.163 ug/ml
Spiked Amount 50.000		Recovery =	58.33%
12) S 1-chlorooctadecane (S...	13.030	3442480	29.549 ug/ml
Spiked Amount 50.000		Recovery =	59.10%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110822AL\
Data File : FC062111.D
Signal(s) : FID1A.ch
Acq On : 08 Nov 2022 18:30
Operator : YP/AJ
Sample : N5506-13
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
TP-10

Integration File: autoint1.e
Quant Time: Nov 09 03:47:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 101722.M
Quant Title : GC Extractables
QLast Update : Mon Oct 17 23:26:31 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.594 min
Delta R.T.: 0.009 min
Response: 4334922
Conc: 29.16 ug/ml

Instrument :
FID_C
ClientSampleId :
TP-10



#12 1-chlorooctadecane (SURR)

R.T.: 13.030 min
Delta R.T.: 0.006 min
Response: 3442480
Conc: 29.55 ug/ml