

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/09/2023
Supervised By :Ankita Jodhani 11/09/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\F0
Data File : FC065071.D
Signal(s) : FID1A.ch
Acq On : 08 Nov 2023 19:46
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.308	3.287	3.389	VB	468899	3852325	88.59%	5.016%
2	4.369	4.325	4.448	BB	422333	3904343	89.78%	5.083%
3	5.949	5.908	6.087	BB	370180	3980892	91.54%	5.183%
4	6.379	6.334	6.462	BB	391733	3937080	90.54%	5.126%
5	7.002	6.958	7.162	BB	316645	3825992	87.98%	4.981%
6	8.166	8.119	8.255	BB	352971	3847533	88.48%	5.009%
7	9.761	9.711	9.848	BB	339740	3829374	88.06%	4.986%
8	11.196	11.149	11.275	BB	324345	3776931	86.85%	4.917%
9	11.508	11.464	11.587	BB	402006	4348611	100.00%	5.662%
10	12.499	12.452	12.564	BB	297537	3564763	81.97%	4.641%
11	12.941	12.889	13.021	BB	259680	3138358	72.17%	4.086%
12	13.107	13.055	13.172	BB	292740	3466276	79.71%	4.513%
13	13.691	13.617	13.754	BB	270052	3386384	77.87%	4.409%
14	14.790	14.737	14.857	BB	250277	3265788	75.10%	4.252%
15	15.807	15.749	15.874	BB	247972	3163821	72.75%	4.119%
16	16.752	16.697	16.811	BB	218273	3110982	71.54%	4.050%
17	17.637	17.583	17.691	BB	239087	3220291	74.05%	4.193%
18	18.465	18.401	18.521	BB	229337	3231147	74.30%	4.207%
19	19.245	19.185	19.304	BB	210595	3120712	71.76%	4.063%
20	19.981	19.950	20.025	BV	209996	2984540	68.63%	3.886%
21	20.702	20.650	20.759	BB	171616	2969936	68.30%	3.867%
22	21.581	21.530	21.684	BB	127260	2881349	66.26%	3.751%
Sum of corrected areas:						76807428		

Aliphatic EPH 102823.M Thu Nov 09 00:10:14 2023