

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110824AL\
 Data File : FC067677.D
 Signal(s) : FID1A.ch
 Acq On : 08 Nov 2024 20:33
 Operator : YP/AJ
 Sample : P4718-02
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Nov 09 01:38:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102324.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 23 18:12:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.929	3231090	29.921 ug/ml
Spiked Amount 50.000		Recovery =	59.84%
Target Compounds			
4) T n-Dodecane (C12)	6.349	48858	0.361 ug/ml
6) T n-Tetradecane (C14)	8.141	47786	0.364 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC110824AL\
Data File : FC067677.D
Signal(s) : FID1A.ch
Acq On : 08 Nov 2024 20:33
Operator : YP/AJ
Sample : P4718-02
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Nov 09 01:38:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102324.M
Quant Title : GC Extractables
QLast Update : Wed Oct 23 18:12:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#4 n-Dodecane (C12)


R.T.: 6.349 min
Delta R.T.: -0.005 min
Response: 48858
Conc: 0.36 ug/ml

Instrument :
FID_C
ClientSampleId :




#6 n-Tetradecane (C14)

R.T.: 8.141 min
Delta R.T.: -0.005 min
Response: 47786
Conc: 0.36 ug/ml



#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.494 min
Response: 0
Conc: N.D.



#12 1-chlorooctadecane (SURR)

R.T.: 12.929 min
Delta R.T.: -0.004 min
Response: 3231090
Conc: 29.92 ug/ml