

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC111224AL\  
 Data File : FC067721.D  
 Signal(s) : FID1A.ch  
 Acq On : 12 Nov 2024 22:49  
 Operator : YP/AJ  
 Sample : P4810-04  
 Misc :  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 FID\_C  
 ClientSampleId :  
 2

Integration File: autoint1.e  
 Quant Time: Nov 13 03:55:46 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 102324.M  
 Quant Title : GC Extractables  
 QLast Update : Wed Oct 23 18:12:58 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response | Conc Units   |
|--------------------------------|--------|----------|--------------|
| -----                          |        |          |              |
| System Monitoring Compounds    |        |          |              |
| 9) S ortho-Terphenyl (SURR)    | 11.487 | 3000573  | 20.449 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 40.90%     |
| 12) S 1-chlorooctadecane (S... | 12.925 | 2097486  | 19.423 ug/ml |
| Spiked Amount                  | 50.000 | Recovery | = 38.85%     |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC111224AL\  
Data File : FC067721.D  
Signal(s) : FID1A.ch  
Acq On : 12 Nov 2024 22:49  
Operator : YP/AJ  
Sample : P4810-04  
Misc :  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :  
2

Integration File: autoint1.e  
Quant Time: Nov 13 03:55:46 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 102324.M  
Quant Title : GC Extractables  
QLast Update : Wed Oct 23 18:12:58 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.487 min  
Delta R.T.: -0.007 min  
Response: 3000573  
Conc: 20.45 ug/ml

Instrument :  
FID\_C  
ClientSampleId :  
2



#12 1-chlorooctadecane (SURR)

R.T.: 12.925 min  
Delta R.T.: -0.008 min  
Response: 2097486  
Conc: 19.42 ug/ml