

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC111318AL\
 Data File : FC037672.D
 Signal(s) : FID1A.CH
 Acq On : 14 Nov 2018 15:33
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 14 15:58:40 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 111318.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 14 03:04:33 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	82.023	88.157 E3	-7.5	100	0.00
2 T	n-Decane (C10)	82.398	89.939 E3	-9.2	101	0.00
3 T	A~Naphthalene (C11.7)	93.305	105.493 E3	-13.1	107	0.00
4 T	n-Dodecane (C12)	84.972	93.193 E3	-9.7	103	0.00
5 T	A~2-methylnaphthalene (C12.	94.996	106.464 E3	-12.1	108	0.00
6 T	n-Tetradecane (C14)	89.423	96.811 E3	-8.3	104	0.00
7 T	n-Hexadecane (C16)	92.769	102.550 E3	-10.5	107	0.00
8 T	n-Octadecane (C18)	98.240	106.909 E3	-8.8	106	0.00
9 S	ortho-Terphenyl (SURR)	109.104	127.915 E3	-17.2	114	0.00
10 T	n-Eicosane (C20)	99.920	110.300 E3	-10.4	107	0.00
11 T	n-Heneicosane (C21)	98.840	110.753 E3	-12.1	108	0.00
12 S	1-chlorooctadecane (SURR)	85.740	95.824 E3	-11.8	108	0.00
13 T	n-Docosane (C22)	100.302	111.586 E3	-11.3	107	0.00
14 T	n-Tetracosane (C24)	101.605	112.347 E3	-10.6	105	0.00
15 T	n-Hexacosane (C26)	101.155	110.312 E3	-9.1	104	0.00
16 T	n-Octacosane (C28)	100.046	108.524 E3	-8.5	105	0.00
17 T	n-Tricontane (C30)	98.208	105.186 E3	-7.1	106	0.00
18 T	n-Dotriacontane (C32)	91.885	97.074 E3	-5.6	105	0.00
19 T	n-Tetratriacontane (C34)	79.981	83.227 E3	-4.1	102	0.00
20 T	n-Hexatriacontane (C36)	71.921	73.050 E3	-1.6	99	0.00
21 T	n-Octatriacontane (C38)	70.366	69.070 E3	1.8	96	0.00
22 T	n-Tetracontane (C40)	71.599	74.144 E3	-3.6	102	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0