

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC111823AL\
 Data File : FC065218.D
 Signal(s) : FID1A.ch
 Acq On : 18 Nov 2023 15:24
 Operator : YP/AJ
 Sample : PB157265BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Nov 20 04:11:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 28 02:12:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.513	9224526	41.914 ug/ml
Spiked Amount 50.000		Recovery =	83.83%
12) S 1-chlorooctadecane (S...	12.945	6860861	43.400 ug/ml
Spiked Amount 50.000		Recovery =	86.80%
Target Compounds			
15) T n-Hexacosane (C26)	15.809	50497	0.315 ug/ml
16) T n-Octacosane (C28)	16.753	129264	0.826 ug/ml
17) T n-Tricontane (C30)	17.636	277406	1.751 ug/ml
18) T n-Dotriacontane (C32)	18.464	338748	2.145 ug/ml
19) T n-Tetratriacontane (C34)	19.245	244010	1.599 ug/ml
20) T n-Hexatriacontane (C36)	19.981	177345	1.175 ug/ml
21) T n-Octatriacontane (C38)	20.702	97119	0.653 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC111823AL\
Data File : FC065218.D
Signal(s) : FID1A.ch
Acq On : 18 Nov 2023 15:24
Operator : YP/AJ
Sample : PB157265BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Nov 20 04:11:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
Quant Title : GC Extractables
QLast Update : Sat Oct 28 02:12:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.513 min
Delta R.T.: -0.010 min
Response: 9224526
Conc: 41.91 ug/ml

Instrument :
FID_C
ClientSampleId :




#12 1-chlorooctadecane (SURR)

R.T.: 12.945 min
Delta R.T.: -0.010 min
Response: 6860861
Conc: 43.40 ug/ml




#15 n-Hexacosane (C26)

R.T.: 15.809 min
Delta R.T.: -0.010 min
Response: 50497
Conc: 0.31 ug/ml



#16 n-Octacosane (C28)


R.T.: 16.753 min
Delta R.T.: -0.011 min
Response: 129264
Conc: 0.83 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.636 min
Delta R.T.: -0.014 min
Response: 277406
Conc: 1.75 ug/ml

Instrument :
FID_C
ClientSampleId :




#18 n-Dotriacontane (C32)

R.T.: 18.464 min
Delta R.T.: -0.013 min
Response: 338748
Conc: 2.15 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.245 min
Delta R.T.: -0.011 min
Response: 244010
Conc: 1.60 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 19.981 min
Delta R.T.: -0.012 min
Response: 177345
Conc: 1.18 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 20.702 min
Delta R.T.: -0.013 min
Response: 97119
Conc: 0.65 ug/ml

Instrument :

FID_C

ClientSampleId :