

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC111918AL\
 Data File : FC037757.D
 Signal(s) : FID1A.CH
 Acq On : 19 Nov 2018 10:31
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 20 01:03:43 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 111318.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 14 03:04:33 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	82.023	83.624 E3	-2.0	95	0.00
2 T	n-Decane (C10)	82.398	89.201 E3	-8.3	100	0.00
3 T	A~Naphthalene (C11.7)	93.305	106.471 E3	-14.1	108	0.00
4 T	n-Dodecane (C12)	84.972	94.788 E3	-11.6	104	0.00
5 T	A~2-methylnaphthalene (C12.	94.996	108.862 E3	-14.6	110	0.00
6 T	n-Tetradecane (C14)	89.423	97.874 E3	-9.5	105	0.00
7 T	n-Hexadecane (C16)	92.769	99.467 E3	-7.2	103	0.00
8 T	n-Octadecane (C18)	98.240	99.404 E3	-1.2	99	0.00
9 S	ortho-Terphenyl (SURR)	109.104	121.766 E3	-11.6	108	0.00
10 T	n-Eicosane (C20)	99.920	98.611 E3	1.3	95	0.00
11 T	n-Heneicosane (C21)	98.840	97.819 E3	1.0	95	0.00
12 S	1-chlorooctadecane (SURR)	85.740	85.167 E3	0.7	96	0.00
13 T	n-Docosane (C22)	100.302	97.913 E3	2.4	94	0.00
14 T	n-Tetracosane (C24)	101.605	98.206 E3	3.3	92	0.00
15 T	n-Hexacosane (C26)	101.155	97.560 E3	3.6	92	0.00
16 T	n-Octacosane (C28)	100.046	99.622 E3	0.4	96	0.00
17 T	n-Tricontane (C30)	98.208	101.470 E3	-3.3	102	0.00
18 T	n-Dotriacontane (C32)	91.885	96.108 E3	-4.6	104	0.00
19 T	n-Tetratriacontane (C34)	79.981	78.709 E3	1.6	97	0.00
20 T	n-Hexatriacontane (C36)	71.921	66.875 E3	7.0	91	0.00
21 T	n-Octatriacontane (C38)	70.366	64.476 E3	8.4	90	0.00
22 T	n-Tetracontane (C40)	71.599	66.967 E3	6.5	92	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0