

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC111918AL\
 Data File : FC037771.D
 Signal(s) : FID1A.CH
 Acq On : 19 Nov 2018 21:25
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 20 01:07:01 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111318.M
 Quant Title : GC Extractables
 QLast Update : Wed Nov 14 03:04:33 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	82.023	94.185 E3	-14.8	107	0.00
2 T	n-Decane (C10)	82.398	100.439 E3	-21.9	113	0.00
3 T	A~Naphthalene (C11.7)	93.305	113.051 E3	-21.2	114	0.00
4 T	n-Dodecane (C12)	84.972	105.618 E3	-24.3	116	0.00
5 T	A~2-methylnaphthalene (C12.	94.996	115.164 E3	-21.2	116	0.00
6 T	n-Tetradecane (C14)	89.423	110.260 E3	-23.3	118	0.00
7 T	n-Hexadecane (C16)	92.769	112.824 E3	-21.6	117	0.00
8 T	n-Octadecane (C18)	98.240	113.160 E3	-15.2	113	0.00
9 S	ortho-Terphenyl (SURR)	109.104	134.715 E3	-23.5	120	0.00
10 T	n-Eicosane (C20)	99.920	113.868 E3	-14.0	110	0.00
11 T	n-Heneicosane (C21)	98.840	113.467 E3	-14.8	110	0.00
12 S	1-chlorooctadecane (SURR)	85.740	98.683 E3	-15.1	111	0.00
13 T	n-Docosane (C22)	100.302	114.114 E3	-13.8	109	0.00
14 T	n-Tetracosane (C24)	101.605	114.890 E3	-13.1	107	0.00
15 T	n-Hexacosane (C26)	101.155	114.882 E3	-13.6	108	0.00
16 T	n-Octacosane (C28)	100.046	117.309 E3	-17.3	113	0.00
17 T	n-Tricontane (C30)	98.208	118.895 E3	-21.1	119	0.00
18 T	n-Dotriacontane (C32)	91.885	111.286 E3	-21.1	121	0.00
19 T	n-Tetratriacontane (C34)	79.981	94.575 E3	-18.2	116	0.00
20 T	n-Hexatriacontane (C36)	71.921	81.183 E3	-12.9	110	0.00
21 T	n-Octatriacontane (C38)	70.366	76.155 E3	-8.2	106	0.00
22 T	n-Tetracontane (C40)	71.599	76.203 E3	-6.4	105	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0