

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/21/2023
Supervised By :Ankita Jodhani 11/21/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\F0
Data File : FC065254.D
Signal(s) : FID1A.ch
Acq On : 21 Nov 2023 07:27
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.306	3.284	3.412	VB	454881	3895272	86.82%	4.996%
2	4.365	4.317	4.487	BB	421654	3954919	88.15%	5.073%
3	5.946	5.909	6.159	BB	352307	4092733	91.22%	5.250%
4	6.375	6.329	6.485	BB	394452	3985706	88.84%	5.112%
5	7.000	6.972	7.192	BB	316744	3880880	86.50%	4.978%
6	8.163	8.117	8.279	BB	373158	3887887	86.66%	4.987%
7	9.756	9.714	9.869	BB	350866	3927078	87.53%	5.037%
8	11.191	11.145	11.300	BB	333700	3888322	86.67%	4.988%
9	11.504	11.452	11.610	BB	419821	4486440	100.00%	5.755%
10	12.493	12.439	12.590	BB	314541	3680914	82.05%	4.722%
11	12.936	12.890	13.034	BV	270312	3230492	72.01%	4.144%
12	13.103	13.034	13.180	VB	292656	3586951	79.95%	4.601%
13	13.686	13.605	13.772	BB	292967	3497925	77.97%	4.487%
14	14.783	14.715	14.867	BB	275209	3354458	74.77%	4.303%
15	15.798	15.742	15.877	BB	249712	3218917	71.75%	4.129%
16	16.743	16.672	16.777	BV	222819	3069817	68.42%	3.938%
17	17.626	17.567	17.700	BB	221408	3185970	71.01%	4.087%
18	18.456	18.400	18.515	BB	225082	3181256	70.91%	4.081%
19	19.236	19.122	19.312	BB	217407	3106078	69.23%	3.984%
20	19.973	19.639	20.016	VB	202651	3008254	67.05%	3.859%
21	20.692	20.650	20.770	BB	168210	2955440	65.87%	3.791%
22	21.565	21.500	21.675	BB	130148	2884719	64.30%	3.700%
Sum of corrected areas:						77960426		

Aliphatic EPH 102823.M Tue Nov 21 08:01:03 2023