

rters

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC11202
Data File : FC067836.D
Signal(s) : FID1A.ch
Acq On : 20 Nov 2024 20:35
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/21/2024
Supervised By :Ankita Jodhani 11/21/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111424.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.287	3.147	3.342	BB	378063	3117239	85.15%	4.697%
2	4.349	4.290	4.419	BB	354212	3207531	87.62%	4.833%
3	5.926	5.869	6.014	BB	371673	3534127	96.54%	5.325%
4	6.363	6.307	6.417	BB	324038	3273242	89.42%	4.932%
5	6.980	6.932	7.090	BB	342191	3446918	94.16%	5.194%
6	8.154	8.095	8.227	BB	296444	3273639	89.43%	4.933%
7	9.751	9.694	9.820	BB	306090	3377517	92.26%	5.089%
8	11.190	11.142	11.272	BB	290922	3416805	93.34%	5.148%
9	11.502	11.434	11.565	BB	328174	3660721	100.00%	5.516%
10	12.497	12.409	12.564	BB	276293	3254924	88.91%	4.904%
11	12.939	12.860	13.002	BB	226010	2741592	74.89%	4.131%
12	13.107	13.037	13.165	BB	250144	3126572	85.41%	4.711%
13	13.692	13.661	13.757	VB	254495	3160142	86.33%	4.762%
14	14.794	14.762	14.850	VB	250406	3210386	87.70%	4.837%
15	15.813	15.740	15.882	BB	229284	3187712	87.08%	4.803%
16	16.761	16.692	16.835	BB	231911	3136239	85.67%	4.726%
17	17.648	17.570	17.707	PB	220463	3089893	84.41%	4.656%
18	18.477	18.410	18.544	BV	198746	2912855	79.57%	4.389%
19	19.259	19.199	19.324	BV	152579	2422773	66.18%	3.651%
20	19.998	19.950	20.031	BV	140040	2044701	55.86%	3.081%
21	20.721	20.590	20.790	BB	109315	1893552	51.73%	2.853%
22	21.607	21.515	21.652	BV	75707	1878343	51.31%	2.830%
Sum of corrected areas:						66367423		

Aliphatic EPH 111424.M Thu Nov 21 02:32:37 2024