

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC112123AL\
 Data File : FC065272.D
 Signal(s) : FID1A.ch
 Acq On : 21 Nov 2023 19:43
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Nov 21 22:52:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 28 02:12:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
17) T n-Tricontane (C30)	17.633	52807	0.333 ug/ml
20) T n-Hexatriacontane (C36)	19.974	2458	0.016 ug/ml
22) T n-Tetracontane (C40)	21.600	105587	0.718 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC112123AL\
Data File : FC065272.D
Signal(s) : FID1A.ch
Acq On : 21 Nov 2023 19:43
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Nov 21 22:52:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
Quant Title : GC Extractables
QLast Update : Sat Oct 28 02:12:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 0.000 min
Exp R.T.: 11.523 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :




#12 1-chlorooctadecane (SURR)

R.T.: 0.000 min
Exp R.T.: 12.955 min
Response: 0
Conc: N.D.




#17 n-Tricontane (C30)

R.T.: 17.633 min
Delta R.T.: -0.017 min
Response: 52807
Conc: 0.33 ug/ml



#20 n-Hexatriacontane (C36)

R.T.: 19.974 min
Delta R.T.: -0.019 min
Response: 2458
Conc: 0.02 ug/ml



#22 n-Tetracontane (C40)

R.T.: 21.600 min
Delta R.T.: 0.002 min
Response: 105587
Conc: 0.72 ug/ml

Instrument :

FID_C

ClientSampleId :