

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/22/2023
Supervised By :Ankita Jodhani 11/22/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\F0
Data File : FC065273.D
Signal(s) : FID1A.ch
Acq On : 21 Nov 2023 20:22
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 102823.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.306	3.284	3.397	VB	462498	3925045	86.04%	4.936%
2	4.366	4.324	4.474	BB	436972	4004238	87.78%	5.036%
3	5.947	5.842	6.119	BB	365451	4101206	89.90%	5.158%
4	6.376	6.332	6.482	BB	402553	4049282	88.77%	5.092%
5	7.001	6.970	7.142	BV	335225	3895023	85.38%	4.898%
6	8.166	8.117	8.274	BB	383344	3964446	86.91%	4.986%
7	9.759	9.664	9.884	BB	361152	3959658	86.80%	4.980%
8	11.192	11.145	11.307	BB	339618	3963900	86.89%	4.985%
9	11.507	11.404	11.602	BB	402728	4561791	100.00%	5.737%
10	12.496	12.440	12.575	BB	315896	3741360	82.02%	4.705%
11	12.938	12.890	13.022	BB	282890	3304347	72.44%	4.155%
12	13.105	13.052	13.187	BB	301944	3643455	79.87%	4.582%
13	13.688	13.609	13.769	BB	298305	3564013	78.13%	4.482%
14	14.786	14.717	14.872	BB	274528	3430145	75.19%	4.314%
15	15.801	15.747	15.875	BB	247238	3287389	72.06%	4.134%
16	16.748	16.665	16.779	BV	230426	3136301	68.75%	3.944%
17	17.630	17.575	17.677	BB	230300	3249807	71.24%	4.087%
18	18.459	18.400	18.502	BB	239286	3250489	71.25%	4.088%
19	19.239	19.164	19.276	PV	224977	3143389	68.91%	3.953%
20	19.976	19.920	20.030	VV	216906	3133117	68.68%	3.940%
21	20.695	20.650	20.743	BV	177486	3025694	66.33%	3.805%
22	21.573	21.500	21.700	BBA	131267	3184376	69.81%	4.005%
Sum of corrected areas:						79518470		

Aliphatic EPH 102823.M Wed Nov 22 03:47:14 2023