

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC112421AL\  
 Data File : FC057070.D  
 Signal(s) : FID1A.CH  
 Acq On : 24 Nov 2021 20:23  
 Operator : AJ\MA  
 Sample : M4814-02DL 5X  
 Misc :  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 FID\_C  
 ClientSampleId :  
 PL-1-112221-EPH-2DL

Integration File: autoint1.e  
 Quant Time: Nov 25 02:08:38 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 111921.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Nov 19 00:52:49 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units  |
|--------------------------------|--------|------------|-------------|
| -----                          |        |            |             |
| System Monitoring Compounds    |        |            |             |
| 9) S ortho-Terphenyl (SURR)    | 11.590 | 644909     | 5.557 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 11.11%      |
| 12) S 1-chlorooctadecane (S... | 13.023 | 552082     | 6.064 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 12.13%      |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC112421AL\  
Data File : FC057070.D  
Signal(s) : FID1A.CH  
Acq On : 24 Nov 2021 20:23  
Operator : AJ\MA  
Sample : M4814-02DL 5X  
Misc :  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :  
PL-1-112221-EPH-2DL

Integration File: autoint1.e  
Quant Time: Nov 25 02:08:38 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Aliphatic EPH 111921.M  
Quant Title : GC Extractables  
QLast Update : Fri Nov 19 00:52:49 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.590 min  
Delta R.T.: -0.006 min  
Response: 644909  
Conc: 5.56 ug/ml

Instrument :  
FID\_C  
ClientSampleId :  
PL-1-112221-EPH-2DL



#12 1-chlorooctadecane (SURR)

R.T.: 13.023 min  
Delta R.T.: -0.006 min  
Response: 552082  
Conc: 6.06 ug/ml