

rters

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Ankita Jodhani 11/29/2021
Supervised By :mohammad ahmed 11/29/2021

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC11262
Data File : FC057098.D
Signal(s) : FID1A.CH
Acq On : 26 Nov 2021 13:30
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111921.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.359	3.331	3.421	BB	281064	2329289	87.55%	4.566%
2	4.425	4.388	4.498	BB	265606	2384750	89.64%	4.675%
3	6.014	5.985	6.115	BB	267872	2660430	100.00%	5.215%
4	6.442	6.408	6.505	BB	248997	2421823	91.03%	4.748%
5	7.071	7.025	7.175	BB	256237	2650113	99.61%	5.195%
6	8.236	8.201	8.308	BB	229913	2423340	91.09%	4.751%
7	9.837	9.801	9.905	BB	224386	2415545	90.80%	4.735%
8	11.277	11.245	11.345	BB	224355	2473310	92.97%	4.848%
9	11.597	11.561	11.658	BB	245425	2575986	96.83%	5.050%
10	12.585	12.545	12.655	BB	200266	2408513	90.53%	4.721%
11	13.030	12.995	13.098	BB	171814	2002532	75.27%	3.926%
12	13.195	13.158	13.265	BB	209327	2403067	90.33%	4.711%
13	13.782	13.711	13.851	BB	202158	2370557	89.10%	4.647%
14	14.882	14.848	14.948	BB	194221	2296112	86.31%	4.501%
15	15.899	15.838	15.965	BB	172822	2213532	83.20%	4.339%
16	16.847	16.778	16.911	BB	169232	2172069	81.64%	4.258%
17	17.730	17.651	17.785	BB	169799	2178811	81.90%	4.271%
18	18.560	18.331	18.601	BB	161683	2163976	81.34%	4.242%
19	19.341	19.141	19.385	BB	153588	2141754	80.50%	4.199%
20	20.076	19.741	20.131	PV	147136	2272650	85.42%	4.455%
21	20.811	20.698	20.861	BB	121619	2040514	76.70%	4.000%
22	21.718	21.628	21.775	BB	83638	2013348	75.68%	3.947%
Sum of corrected areas:						51012022		

Aliphatic EPH 111921.M Fri Nov 26 13:57:45 2021