

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112921\
 Data File : PR052702.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2021 20:11
 Operator : AJ\MA
 Sample : PR112921ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 04:05:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 29 22:45:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	3.296	3.588 E6	-8.9	101	0.00
2 SA Decachlorobiphenyl	3.241	3.381 E6	-4.3	104	0.00
3 L1 AR-1016-1	123.603	130.383 E3	-5.5	101	0.00
4 L1 AR-1016-2	173.159	182.922 E3	-5.6	101	0.00
5 L1 AR-1016-3	108.467	114.039 E3	-5.1	101	0.00
6 L1 AR-1016-4	88.509	94.973 E3	-7.3	102	0.00
7 L1 AR-1016-5	90.133	95.604 E3	-6.1	101	0.00
31 L7 AR-1260-1	155.299	161.548 E3	-4.0	101	0.00
32 L7 AR-1260-2	186.597	193.490 E3	-3.7	102	0.00
33 L7 AR-1260-3	146.543	151.745 E3	-3.5	101	0.00
34 L7 AR-1260-4	170.439	176.274 E3	-3.4	101	0.00
35 L7 AR-1260-5	332.076	352.033 E3	-6.0	103	0.00

Signal #2

1 SA Tetrachloro-m-xylene	2.730	2.933 E6	-7.4	101	0.00
2 SA Decachlorobiphenyl	3.232	3.301 E6	-2.1	100	0.00
3 L1 AR-1016-1	97.183	101.217 E3	-4.2	101	0.00
4 L1 AR-1016-2	145.007	153.510 E3	-5.9	101	0.00
5 L1 AR-1016-3	77.385	82.590 E3	-6.7	101	0.00
6 L1 AR-1016-4	63.327	65.793 E3	-3.9	101	0.00
7 L1 AR-1016-5	80.979	84.249 E3	-4.0	100	0.00
31 L7 AR-1260-1	142.319	147.148 E3	-3.4	101	0.00
32 L7 AR-1260-2	173.610	181.271 E3	-4.4	102	0.00
33 L7 AR-1260-3	168.722	175.932 E3	-4.3	101	0.00
34 L7 AR-1260-4	151.883	157.058 E3	-3.4	101	0.00
35 L7 AR-1260-5	366.316	386.688 E3	-5.6	101	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	40.842	0.000 E3	100.0#	0#	-4.90#
9 L2 AR-1221-2	33.681	0.000 E3	100.0#	0#	-4.99#
10 L2 AR-1221-3	92.161	0.000 E3	100.0#	0#	-5.07#
11 L3 AR-1232-1	78.207	0.000 E3	100.0#	0#	-5.07#
12 L3 AR-1232-2	41.781	0.000 E3	100.0#	0#	-5.60#
13 L3 AR-1232-3	76.693	0.000 E3	100.0#	0#	-5.90#
14 L3 AR-1232-4	39.569	0.000 E3	100.0#	0#	-6.06#
15 L3 AR-1232-5	31.729	0.000 E3	100.0#	0#	-6.14#
16 L4 AR-1242-1	95.847	0.000 E3	100.0#	0#	-5.87#
17 L4 AR-1242-2	129.119	0.000 E3	100.0#	0#	-5.90#
18 L4 AR-1242-3	82.785	0.000 E3	100.0#	0#	-5.96#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112921\
 Data File : PR052702.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2021 20:11
 Operator : AJ\MA
 Sample : PR112921ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 04:05:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 29 22:45:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	68.179	0.000 E3	100.0#	0#	-6.06#
20 L4 AR-1242-5	74.474	0.000 E3	100.0#	0#	-6.80#
21 L5 AR-1248-1	71.271	0.000 E3	100.0#	0#	-5.87#
22 L5 AR-1248-2	101.550	0.000 E3	100.0#	0#	-6.15#
23 L5 AR-1248-3	112.676	0.000 E3	100.0#	0#	-6.36#
24 L5 AR-1248-4	126.676	0.000 E3	100.0#	0#	-6.76#
25 L5 AR-1248-5	122.022	0.000 E3	100.0#	0#	-6.80#
26 L6 AR-1254-1	131.247	0.000 E3	100.0#	0#	-6.73#
27 L6 AR-1254-2	202.323	0.000 E3	100.0#	0#	-6.95#
28 L6 AR-1254-3	212.004	0.000 E3	100.0#	0#	-7.32#
29 L6 AR-1254-4	157.922	0.000 E3	100.0#	0#	-7.61#
30 L6 AR-1254-5	176.244	0.000 E3	100.0#	0#	-8.03#
36 L8 AR-1262-1	202.791	0.000 E3	100.0#	0#	-8.10#
37 L8 AR-1262-2	365.607	0.000 E3	100.0#	0#	-8.68#
38 L8 AR-1262-3	163.715	0.000 E3	100.0#	0#	-9.01#
39 L8 AR-1262-4	114.262	0.000 E3	100.0#	0#	-9.10#
40 L8 AR-1262-5	139.883	0.000 E3	100.0#	0#	-9.82#
41 L9 AR-1268-1	447.169	0.000 E3	100.0#	0#	-9.01#
42 L9 AR-1268-2	411.145	0.000 E3	100.0#	0#	-9.11#
43 L9 AR-1268-3	352.562	0.000 E3	100.0#	0#	-9.35#
44 L9 AR-1268-4	159.446	0.000 E3	100.0#	0#	-9.82#
45 L9 AR-1268-5	1.161	0.000 E6	100.0#	0#	-10.26#

Signal #2

8 L2 AR-1221-1	34.872	0.000 E3	100.0#	0#	-4.07#
9 L2 AR-1221-2	25.713	0.000 E3	100.0#	0#	-4.15#
10 L2 AR-1221-3	79.584	0.000 E3	100.0#	0#	-4.23#
11 L3 AR-1232-1	66.780	0.000 E3	100.0#	0#	-4.23#
12 L3 AR-1232-2	64.006	0.000 E3	100.0#	0#	-4.95#
13 L3 AR-1232-3	34.414	0.000 E3	100.0#	0#	-5.13#
14 L3 AR-1232-4	30.894	0.000 E3	100.0#	0#	-5.21#
15 L3 AR-1232-5	32.955	0.000 E3	100.0#	0#	-5.38#
16 L4 AR-1242-1	72.238	0.000 E3	100.0#	0#	-4.93#
17 L4 AR-1242-2	109.164	0.000 E3	100.0#	0#	-4.95#
18 L4 AR-1242-3	59.292	0.000 E3	100.0#	0#	-5.13#
19 L4 AR-1242-4	57.893	0.000 E3	100.0#	0#	-5.21#
20 L4 AR-1242-5	70.685	0.000 E3	100.0#	0#	-5.73#
21 L5 AR-1248-1	57.275	0.000 E3	100.0#	0#	-4.93#
22 L5 AR-1248-2	79.414	0.000 E3	100.0#	0#	-5.17#
23 L5 AR-1248-3	84.301	0.000 E3	100.0#	0#	-5.21#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112921\
 Data File : PR052702.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2021 20:11
 Operator : AJ\MA
 Sample : PR112921ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 04:05:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 29 22:45:34 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	99.169	0.000 E3	100.0#	0#	-5.38#
25 L5 AR-1248-5	97.699	0.000 E3	100.0#	0#	-5.77#
26 L6 AR-1254-1	151.615	0.000 E3	100.0#	0#	-5.73#
27 L6 AR-1254-2	131.139	0.000 E3	100.0#	0#	-5.87#
28 L6 AR-1254-3	220.745	0.000 E3	100.0#	0#	-6.28#
29 L6 AR-1254-4	136.220	0.000 E3	100.0#	0#	-6.50#
30 L6 AR-1254-5	207.804	0.000 E3	100.0#	0#	-6.92#
36 L8 AR-1262-1	108.335	0.000 E3	100.0#	0#	-6.92#
37 L8 AR-1262-2	388.931	0.000 E3	100.0#	0#	-7.45#
38 L8 AR-1262-3	159.481	0.000 E3	100.0#	0#	-7.73#
39 L8 AR-1262-4	301.774	0.000 E3	100.0#	0#	-7.80#
40 L8 AR-1262-5	144.505	0.000 E3	100.0#	0#	-8.30#
41 L9 AR-1268-1	472.919	0.000 E3	100.0#	0#	-7.73#
42 L9 AR-1268-2	445.271	0.000 E3	100.0#	0#	-7.80#
43 L9 AR-1268-3	386.111	0.000 E3	100.0#	0#	-8.01#
44 L9 AR-1268-4	163.064	0.000 E3	100.0#	0#	-8.30#
45 L9 AR-1268-5	1.247	0.000 E6	100.0#	0#	-8.60#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0