

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120122AL\
 Data File : FC062338.D
 Signal(s) : FID1A.ch
 Acq On : 01 Dec 2022 09:52
 Operator : YP/AJ
 Sample : N5812-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Dec 01 12:30:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111422.M
 Quant Title : GC Extractables
 QLast Update : Mon Nov 21 12:12:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.593	4346331	31.842 ug/ml
Spiked Amount 50.000		Recovery =	63.68%
12) S 1-chlorooctadecane (S...	13.028	3183923	31.322 ug/ml
Spiked Amount 50.000		Recovery =	62.64%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120122AL\
Data File : FC062338.D
Signal(s) : FID1A.ch
Acq On : 01 Dec 2022 09:52
Operator : YP/AJ
Sample : N5812-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Dec 01 12:30:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111422.M
Quant Title : GC Extractables
QLast Update : Mon Nov 21 12:12:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.593 min
Delta R.T.: -0.012 min
Response: 4346331
Conc: 31.84 ug/ml

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 13.028 min
Delta R.T.: -0.014 min
Response: 3183923
Conc: 31.32 ug/ml