

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120122AL\
 Data File : FC062358.D
 Signal(s) : FID1A.ch
 Acq On : 02 Dec 2022 11:45
 Operator : YP/AJ
 Sample : N5806-03
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Dec 02 15:17:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111422.M
 Quant Title : GC Extractables
 QLast Update : Mon Nov 21 12:12:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.585	483330	3.541 ug/ml
Spiked Amount 50.000		Recovery =	7.08%
12) S 1-chlorooctadecane (S...	13.026	3220878	31.686 ug/ml
Spiked Amount 50.000		Recovery =	63.37%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120122AL\
Data File : FC062358.D
Signal(s) : FID1A.ch
Acq On : 02 Dec 2022 11:45
Operator : YP/AJ
Sample : N5806-03
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Dec 02 15:17:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111422.M
Quant Title : GC Extractables
QLast Update : Mon Nov 21 12:12:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.585 min
Delta R.T.: -0.020 min
Response: 483330
Conc: 3.54 ug/ml

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 13.026 min
Delta R.T.: -0.016 min
Response: 3220878
Conc: 31.69 ug/ml