

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120122AL\
Data File : FC062361.D
Signal(s) : FID1A.ch
Acq On : 02 Dec 2022 13:47
Operator : YP/AJ
Sample : N5806-10
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Dec 02 15:18:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111422.M
Quant Title : GC Extractables
QLast Update : Mon Nov 21 12:12:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	13.023	2393282	23.544 ug/ml
Spiked Amount 50.000		Recovery =	47.09%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120122AL\
Data File : FC062361.D
Signal(s) : FID1A.ch
Acq On : 02 Dec 2022 13:47
Operator : YP/AJ
Sample : N5806-10
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Dec 02 15:18:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111422.M
Quant Title : GC Extractables
QLast Update : Mon Nov 21 12:12:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.605 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 13.023 min
Delta R.T.: -0.019 min
Response: 2393282
Conc: 23.54 ug/ml