

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120320AL\
 Data File : FC051136.D
 Signal(s) : FID1A.CH
 Acq On : 04 Dec 2020 16:38
 Operator : DD\AJ
 Sample : L4916-01DL 10X
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 OK-01-12220DL

Manual Integrations
 APPROVED

Ankita
 12/7/2020 12:16:40 PM

Integration File: autoint1.e
 Quant Time: Dec 04 22:04:20 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 112720.M
 Quant Title : GC Extractables
 QLast Update : Mon Nov 30 10:42:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.591	287671	2.338 ug/ml
Spiked Amount	50.000	Recovery	= 4.68%
12) S 1-chlorooctadecane (S...	13.025	224931	2.306 ug/mlm
Spiked Amount	50.000	Recovery	= 4.61%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120320AL\
Data File : FC051136.D
Signal(s) : FID1A.CH
Acq On : 04 Dec 2020 16:38
Operator : DD\AJ
Sample : L4916-01DL 10X
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
OK-01-12220DL


Manual Integrations
APPROVED

Ankita
12/7/2020 12:16:40 PM

Integration File: autoint1.e
Quant Time: Dec 04 22:04:20 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 112720.M
Quant Title : GC Extractables
QLast Update : Mon Nov 30 10:42:25 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.591 min
Delta R.T.: -0.009 min
Response: 287671
Conc: 2.34 ug/ml

Instrument :
FID_C
ClientSampleId :
OK-01-12220DL

Manual Integrations
APPROVED

Ankita
12/7/2020 12:16:40 PM



#12 1-chlorooctadecane (SURR)

R.T.: 13.025 min
Delta R.T.: -0.008 min
Response: 224931
Conc: 2.31 ug/ml m