

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120924AL\
 Data File : FC067953.D
 Signal(s) : FID1A.ch
 Acq On : 09 Dec 2024 18:07
 Operator : YP/AJ
 Sample : PB165500BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Dec 10 04:41:26 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111424.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 14 14:13:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.913	5644971	38.888 ug/ml
Spiked Amount 50.000		Recovery =	77.78%
Target Compounds			
4) T n-Dodecane (C12)	6.329	69822	0.415 ug/ml
6) T n-Tetradecane (C14)	8.121	50001	0.299 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC120924AL\
Data File : FC067953.D
Signal(s) : FID1A.ch
Acq On : 09 Dec 2024 18:07
Operator : YP/AJ
Sample : PB165500BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Dec 10 04:41:26 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111424.M
Quant Title : GC Extractables
QLast Update : Thu Nov 14 14:13:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#4 n-Dodecane (C12)


R.T.: 6.329 min
Delta R.T.: -0.019 min
Response: 69822
Conc: 0.41 ug/ml

Instrument :
FID_C
ClientSampleId :




#6 n-Tetradecane (C14)

R.T.: 8.121 min
Delta R.T.: -0.018 min
Response: 50001
Conc: 0.30 ug/ml



#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T. : 11.486 min
Response: 0
Conc: N.D.



#12 1-chlorooctadecane (SURR)

R.T.: 12.913 min
Delta R.T.: -0.012 min
Response: 5644971
Conc: 38.89 ug/ml