

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC121218AL\
 Data File : FC038154.D
 Signal(s) : FID1A.CH
 Acq On : 12 Dec 2018 11:24
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: sample.E
 Quant Time: Dec 12 11:53:12 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 121018.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 11 15:29:14 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	103.997	111.121 E3	-6.9	109	0.00
2 T	n-Decane (C10)	104.733	112.537 E3	-7.5	109	0.00
3 T	A~Naphthalene (C11.7)	117.468	125.536 E3	-6.9	108	0.00
4 T	n-Dodecane (C12)	107.961	114.671 E3	-6.2	108	0.00
5 T	A~2-methylnaphthalene (C12.	117.500	124.022 E3	-5.6	106	0.00
6 T	n-Tetradecane (C14)	111.691	115.413 E3	-3.3	105	0.00
7 T	n-Hexadecane (C16)	115.362	115.974 E3	-0.5	102	0.00
8 T	n-Octadecane (C18)	120.880	116.677 E3	3.5	99	0.00
9 S	ortho-Terphenyl (SURR)	130.366	128.786 E3	1.2	100	0.00
10 T	n-Eicosane (C20)	120.111	115.985 E3	3.4	98	0.00
11 T	n-Heneicosane (C21)	116.037	115.474 E3	0.5	101	0.00
12 S	1-chlorooctadecane (SURR)	100.823	98.655 E3	2.2	99	0.00
13 T	n-Docosane (C22)	115.573	115.653 E3	-0.1	101	0.00
14 T	n-Tetracosane (C24)	114.815	116.235 E3	-1.2	103	0.00
15 T	n-Hexacosane (C26)	112.490	116.334 E3	-3.4	105	0.00
16 T	n-Octacosane (C28)	112.401	118.405 E3	-5.3	108	0.00
17 T	n-Tricontane (C30)	109.750	119.808 E3	-9.2	113	0.00
18 T	n-Dotriacontane (C32)	102.424	114.727 E3	-12.0	117	0.00
19 T	n-Tetratriacontane (C34)	88.127	100.086 E3	-13.6	118	0.00
20 T	n-Hexatriacontane (C36)	78.343	88.677 E3	-13.2	116	0.00
21 T	n-Octatriacontane (C38)	77.382	86.865 E3	-12.3	117	0.00
22 T	n-Tetracontane (C40)	87.385	91.536 E3	-4.8	112	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0