

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC121223AL\
 Data File : FC065390.D
 Signal(s) : FID1A.ch
 Acq On : 12 Dec 2023 16:50
 Operator : YP/AJ
 Sample : 05760-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 CONCRETE-1

Integration File: autoint1.e
 Quant Time: Dec 12 22:35:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 112423.M
 Quant Title : GC Extractables
 QLast Update : Fri Nov 24 13:28:29 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.504	9496297	40.932 ug/ml
Spiked Amount	50.000	Recovery	= 81.86%
12) S 1-chlorooctadecane (S...	12.935	6929448	41.848 ug/ml
Spiked Amount	50.000	Recovery	= 83.70%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC121223AL\
Data File : FC065390.D
Signal(s) : FID1A.ch
Acq On : 12 Dec 2023 16:50
Operator : YP/AJ
Sample : 05760-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
CONCRETE-1

Integration File: autoint1.e
Quant Time: Dec 12 22:35:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 112423.M
Quant Title : GC Extractables
QLast Update : Fri Nov 24 13:28:29 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

