

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC121318AL\
 Data File : FC038178.D
 Signal(s) : FID1A.CH
 Acq On : 13 Dec 2018 2:54
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: sample.E
 Quant Time: Dec 13 07:39:51 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 121018.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 11 15:29:14 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	103.997	102.016 E3	1.9	100	0.00
2 T	n-Decane (C10)	104.733	102.982 E3	1.7	100	0.00
3 T	A~Naphthalene (C11.7)	117.468	115.807 E3	1.4	99	0.00
4 T	n-Dodecane (C12)	107.961	104.558 E3	3.2	99	0.00
5 T	A~2-methylnaphthalene (C12.	117.500	114.402 E3	2.6	97	0.00
6 T	n-Tetradecane (C14)	111.691	105.676 E3	5.4	97	0.00
7 T	n-Hexadecane (C16)	115.362	106.852 E3	7.4	94	0.00
8 T	n-Octadecane (C18)	120.880	106.965 E3	11.5	91	0.00
9 S	ortho-Terphenyl (SURR)	130.366	121.691 E3	6.7	94	0.00
10 T	n-Eicosane (C20)	120.111	107.897 E3	10.2	91	0.00
11 T	n-Heneicosane (C21)	116.037	107.928 E3	7.0	94	0.00
12 S	1-chlorooctadecane (SURR)	100.823	91.976 E3	8.8	92	0.00
13 T	n-Docosane (C22)	115.573	108.293 E3	6.3	95	0.00
14 T	n-Tetracosane (C24)	114.815	109.667 E3	4.5	97	0.00
15 T	n-Hexacosane (C26)	112.490	109.498 E3	2.7	99	0.00
16 T	n-Octacosane (C28)	112.401	112.550 E3	-0.1	103	0.00
17 T	n-Tricontane (C30)	109.750	114.150 E3	-4.0	107	0.00
18 T	n-Dotriacontane (C32)	102.424	110.020 E3	-7.4	113	0.00
19 T	n-Tetratriacontane (C34)	88.127	95.093 E3	-7.9	112	0.00
20 T	n-Hexatriacontane (C36)	78.343	85.317 E3	-8.9	112	0.00
21 T	n-Octatriacontane (C38)	77.382	81.411 E3	-5.2	110	-0.01
22 T	n-Tetracontane (C40)	87.385	81.414 E3	6.8	99	-0.01

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0