

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC121323AL\
 Data File : FC065426.D
 Signal(s) : FID1A.ch
 Acq On : 13 Dec 2023 20:28
 Operator : YP/AJ
 Sample : PB157747BL
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Dec 13 21:19:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 112423.M
 Quant Title : GC Extractables
 QLast Update : Fri Nov 24 13:28:29 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.502	10931636	47.118 ug/ml
Spiked Amount 50.000		Recovery =	94.24%
12) S 1-chlorooctadecane (S...	12.932	7836222	47.324 ug/ml
Spiked Amount 50.000		Recovery =	94.65%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC121323AL\
Data File : FC065426.D
Signal(s) : FID1A.ch
Acq On : 13 Dec 2023 20:28
Operator : YP/AJ
Sample : PB157747BL
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Dec 13 21:19:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 112423.M
Quant Title : GC Extractables
QLast Update : Fri Nov 24 13:28:29 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.502 min
Delta R.T.: 0.000 min
Response: 10931636
Conc: 47.12 ug/ml

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 12.932 min
Delta R.T.: -0.004 min
Response: 7836222
Conc: 47.32 ug/ml