

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC123022AL\
 Data File : FC062521.D
 Signal(s) : FID1A.ch
 Acq On : 30 Dec 2022 09:40
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Dec 30 14:59:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 120722.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 08 01:05:24 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.551	2499032	17.716 ug/ml
Spiked Amount 50.000		Recovery =	35.43%
12) S 1-chlorooctadecane (S...	12.990	1894908	17.779 ug/ml
Spiked Amount 50.000		Recovery =	35.56%
Target Compounds			
1) T n-Nonane (C9)	3.312	2169908	16.871 ug/ml
2) T n-Decane (C10)	4.379	2205063	17.726 ug/ml
3) T A~Naphthalene (C11.7)	5.964	2492529	17.565 ug/ml
4) T n-Dodecane (C12)	6.400	2157934	17.228 ug/ml
5) T A~2-methylnaphthalene...	7.022	2327200	18.099 ug/ml
6) T n-Tetradecane (C14)	8.196	2213868	18.179 ug/ml
7) T n-Hexadecane (C16)	9.797	2252910	17.949 ug/ml
8) T n-Octadecane (C18)	11.237	2307306	18.061 ug/ml
10) T n-Eicosane (C20)	12.546	2257644	18.149 ug/ml
11) T n-Heneicosane (C21)	13.159	2214399	17.919 ug/ml
13) T n-Docosane (C22)	13.743	2293905	18.884 ug/ml
14) T n-Tetracosane (C24)	14.845	2418311	20.175 ug/ml
15) T n-Hexacosane (C26)	15.864	2475472	21.310 ug/ml
16) T n-Octacosane (C28)	16.814	2587302	22.765 ug/ml
17) T n-Tricontane (C30)	17.699	2672262	23.902 ug/ml
18) T n-Dotriacontane (C32)	18.528	2644612	23.800 ug/ml
19) T n-Tetratriacontane (C34)	19.309	2340521	21.635 ug/ml
20) T n-Hexatriacontane (C36)	20.049	2118495	20.867 ug/ml
21) T n-Octatriacontane (C38)	20.779	2025355	19.796 ug/ml
22) T n-Tetracontane (C40)	21.681	1996396	19.699 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC123022AL\
Data File : FC062521.D
Signal(s) : FID1A.ch
Acq On : 30 Dec 2022 09:40
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
Quant Time: Dec 30 14:59:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 120722.M
Quant Title : GC Extractables
QLast Update : Thu Dec 08 01:05:24 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

