

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC071918AL\
 Data File : FC035523.D
 Signal(s) : FID1A.CH
 Acq On : 20 Jul 2018 15:34
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: sample.E
 Quant Time: Jul 20 16:10:58 2018
 Quant Method : Z:\PESTPCBSRV\HPCHEM1\FID_C\Method\Alaphatic EPH 062518.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 25 14:27:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	83.857	79.891 E3	4.7	92	-0.02
2 T	n-Decane (C10)	83.813	83.095 E3	0.9	96	-0.01
3 T	A-Naphthalene (C11.7)	94.890	104.630 E3	-10.3	107	-0.02
4 T	n-Dodecane (C12)	86.311	89.409 E3	-3.6	102	-0.01
5 T	A-2-methylnaphthalene (C12.)	96.295	106.840 E3	-11.0	109	-0.02
6 T	n-Tetradecane (C14)	89.174	93.043 E3	-4.3	104	-0.01
7 T	n-Hexadecane (C16)	91.744	94.640 E3	-3.2	105	-0.01
8 T	n-Octadecane (C18)	94.376	94.359 E3	0.0	101	-0.01
9 S	ortho-Terphenyl (SURR)	102.960	110.059 E3	-6.9	109	-0.02
10 T	n-Eicosane (C20)	96.430	96.355 E3	0.1	102	-0.02
11 T	n-Heneicosane (C21)	94.999	96.423 E3	-1.5	104	-0.02
12 S	1-chlorooctadecane (SURR)	80.043	81.222 E3	-1.5	103	-0.02
13 T	n-Docosane (C22)	95.442	96.927 E3	-1.6	104	-0.02
14 T	n-Tetracosane (C24)	95.452	95.948 E3	-0.5	102	-0.02
15 T	n-Hexacosane (C26)	94.536	96.295 E3	-1.9	103	-0.02
16 T	n-Octacosane (C28)	94.019	98.331 E3	-4.6	105	-0.02
17 T	n-Tricontane (C30)	91.885	97.954 E3	-6.6	107	-0.02
18 T	n-Dotriacontane (C32)	88.793	98.328 E3	-10.7	111	-0.02
19 T	n-Tetratriacontane (C34)	90.481	93.732 E3	-3.6	104	-0.02
20 T	n-Hexatriacontane (C36)	89.661	89.211 E3	0.5	99	-0.02
21 T	n-Octatriacontane (C38)	92.226	89.544 E3	2.9	99	-0.02
22 T	n-Tetracontane (C40)	93.982	85.252 E3	9.3	91	-0.02

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0