

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD010521AR\
 Data File : FD035889.D
 Signal(s) : FID2B.CH
 Acq On : 05 Jan 2021 14:24
 Operator : DD\AJ
 Sample : L5246-05DL2 20X
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 FID_D
 ClientSampleId :
 5DL2

Manual Integrations
 APPROVED

Ankita
 1/6/2021 1:15:43 PM

Integration File: autoint1.e
 Quant Time: Jan 06 08:22:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 122120.M
 Quant Title : GC Extractables
 QLast Update : Tue Dec 22 09:21:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
4) S 2-Bromonaphthalene (S...	7.588	325595	2.984 ug/mlm
Spiked Amount 50.000		Recovery =	5.97%
6) S 2-Fluorobiphenyl (SURR)	8.500	220109	3.170 ug/ml
Spiked Amount 50.000 Range 0 - 131		Recovery =	6.34%
11) S ortho-Terphenyl (SURR)	11.476	349175	2.701 ug/ml
Spiked Amount 50.000		Recovery =	5.40%
Target Compounds			

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD010521AR\
Data File : FD035889.D
Signal(s) : FID2B.CH
Acq On : 05 Jan 2021 14:24
Operator : DD\AJ
Sample : L5246-05DL2 20X
Misc :
ALS Vial : 59 Sample Multiplier: 1


Instrument :
FID_D
ClientSampleId :
5DL2


Manual Integrations
APPROVED

Ankita
1/6/2021 1:15:43 PM

Integration File: autoint1.e
Quant Time: Jan 06 08:22:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aromatic EPH 122120.M
Quant Title : GC Extractables
QLast Update : Tue Dec 22 09:21:33 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm






#4 2-Bromonaphthalene (SURR)

R.T.: 7.588 min
Delta R.T.: -0.008 min
Response: 325595
Conc: 2.98 ug/ml m

Instrument :
FID_D
ClientSampleId :
5DL2


Manual Integrations
APPROVED

Ankita
1/6/2021 1:15:43 PM



#6 2-Fluorobiphenyl (SURR)

R.T.: 8.500 min
Delta R.T.: 0.028 min
Response: 220109
Conc: 3.17 ug/ml



#11 ortho-Terphenyl (SURR)

R.T.: 11.476 min
Delta R.T.: -0.011 min
Response: 349175
Conc: 2.70 ug/ml