

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD010722AL\
 Data File : FD040497.D
 Signal(s) : FID2B.CH
 Acq On : 08 Jan 2022 2:02
 Operator : AJ\MA
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Jan 08 02:52:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 010422.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 05 07:08:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	128.382	126.006 E3	1.9	96	0.00
2 T n-Decane (C10)	128.766	128.439 E3	0.3	98	0.00
3 T A~Naphthalene (C11.7)	142.072	143.241 E3	-0.8	98	0.00
4 T n-Dodecane (C12)	127.814	130.454 E3	-2.1	100	0.00
5 T A~2-methylnaphthalene (C12.	141.074	144.778 E3	-2.6	100	0.00
6 T n-Tetradecane (C14)	129.540	134.227 E3	-3.6	101	0.00
7 T n-Hexadecane (C16)	130.284	135.662 E3	-4.1	102	0.00
8 T n-Octadecane (C18)	135.501	138.700 E3	-2.4	100	0.00
9 S ortho-Terphenyl (SURR)	154.188	153.056 E3	0.7	98	0.00
10 T n-Eicosane (C20)	136.020	136.422 E3	-0.3	99	0.00
11 T n-Heneicosane (C21)	138.476	136.061 E3	1.7	96	0.00
12 S 1-chlorooctadecane (SURR)	117.707	116.328 E3	1.2	97	0.00
13 T n-Docosane (C22)	137.887	136.220 E3	1.2	97	0.00
14 T n-Tetracosane (C24)	139.825	137.126 E3	1.9	96	0.00
15 T n-Hexacosane (C26)	137.845	133.302 E3	3.3	94	0.00
16 T n-Octacosane (C28)	135.686	131.556 E3	3.0	94	-0.01
17 T n-Tricontane (C30)	130.338	127.229 E3	2.4	96	0.00
18 T n-Dotriacontane (C32)	119.545	116.901 E3	2.2	95	0.00
19 T n-Tetratriacontane (C34)	105.292	105.672 E3	-0.4	99	0.00
20 T n-Hexatriacontane (C36)	101.331	96.001 E3	5.3	96	0.00
21 T n-Octatriacontane (C38)	89.345	89.428 E3	-0.1	98	0.00
22 T n-Tetracontane (C40)	85.482	94.300 E3	-10.3	106	-0.01

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0