

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_D\Data\FD011022AL\
 Data File : FD040504.D
 Signal(s) : FID2B.CH
 Acq On : 10 Jan 2022 12:26
 Operator : AJ\MA
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_D
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Jan 11 02:07:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 010422.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 05 07:08:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	128.382	132.308 E3	-3.1	101	0.00
2 T n-Decane (C10)	128.766	134.692 E3	-4.6	102	0.00
3 T A~Naphthalene (C11.7)	142.072	150.260 E3	-5.8	103	0.00
4 T n-Dodecane (C12)	127.814	136.899 E3	-7.1	105	0.00
5 T A~2-methylnaphthalene (C12.	141.074	151.884 E3	-7.7	105	0.00
6 T n-Tetradecane (C14)	129.540	140.817 E3	-8.7	106	0.00
7 T n-Hexadecane (C16)	130.284	142.639 E3	-9.5	107	0.00
8 T n-Octadecane (C18)	135.501	146.550 E3	-8.2	106	0.00
9 S ortho-Terphenyl (SURR)	154.188	161.544 E3	-4.8	103	0.00
10 T n-Eicosane (C20)	136.020	145.565 E3	-7.0	105	0.00
11 T n-Heneicosane (C21)	138.476	144.888 E3	-4.6	103	0.00
12 S 1-chlorooctadecane (SURR)	117.707	123.045 E3	-4.5	102	0.00
13 T n-Docosane (C22)	137.887	148.631 E3	-7.8	105	0.00
14 T n-Tetracosane (C24)	139.825	156.117 E3	-11.7	109	0.00
15 T n-Hexacosane (C26)	137.845	155.074 E3	-12.5	110	0.00
16 T n-Octacosane (C28)	135.686	154.078 E3	-13.6	111	0.00
17 T n-Tricontane (C30)	130.338	148.158 E3	-13.7	111	0.00
18 T n-Dotriacontane (C32)	119.545	133.744 E3	-11.9	109	0.00
19 T n-Tetratriacontane (C34)	105.292	117.911 E3	-12.0	111	0.00
20 T n-Hexatriacontane (C36)	101.331	105.275 E3	-3.9	105	0.00
21 T n-Octatriacontane (C38)	89.345	98.155 E3	-9.9	107	-0.01
22 T n-Tetracontane (C40)	85.482	98.544 E3	-15.3	111	-0.01

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0